

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082024\
 Data File : VY019149.D
 Acq On : 20 Aug 2024 14:28
 Operator : SY/MD
 Sample : P3634-01RE
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-06-08152024RE

Quant Time: Aug 20 15:14:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	35304	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.603	114	54589	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	38399	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	12325	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	12310	34.784	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	69.560%
35) Dibromofluoromethane	7.622	113	16768	46.978	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.960%
50) Toluene-d8	10.103	98	56742	44.175	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.340%
62) 4-Bromofluorobenzene	12.408	95	13249	30.257	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	60.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082024\
Data File : VY019149.D
Acq On : 20 Aug 2024 14:28
Operator : SY/MD
Sample : P3634-01RE
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-06-08152024RE

Quant Time: Aug 20 15:14:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

