

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082124\
 Data File : VY019179.D
 Acq On : 21 Aug 2024 19:07
 Operator : SY/MD
 Sample : P3643-07RE
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 918-J-SD-0-0.5-081524RE

Quant Time: Aug 22 01:17:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	16538	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	24865	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	21220	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	5869	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	8889	53.619	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.240%	
35) Dibromofluoromethane	7.628	113	9184	56.488	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 112.980%	
50) Toluene-d8	10.103	98	27659	47.274	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 94.540%	
62) 4-Bromofluorobenzene	12.408	95	7215	36.174	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 72.340%	
Target Compounds						
16) Acetone	3.866	43	809	26.314	ug/l	Qvalue # 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082124\
Data File : VY019179.D
Acq On : 21 Aug 2024 19:07
Operator : SY/MD
Sample : P3643-07RE
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
918-J-SD-0-0.5-081524RE

Quant Time: Aug 22 01:17:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

