

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082124\
 Data File : VY019186.D
 Acq On : 21 Aug 2024 21:55
 Operator : SY/MD
 Sample : P3534-03RE
 Misc : 12.37g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HP-IPRE

Quant Time: Aug 22 01:20:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	94621	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	174143	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	131716	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	47633	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	52406	55.251	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.500%
35) Dibromofluoromethane	7.628	113	56236	49.388	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.780%
50) Toluene-d8	10.103	98	168554	41.135	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	82.260%
62) 4-Bromofluorobenzene	12.408	95	51013	36.520	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	73.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082124\
Data File : VY019186.D
Acq On : 21 Aug 2024 21:55
Operator : SY/MD
Sample : P3534-03RE
Misc : 12.37g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HP-IPRE

Quant Time: Aug 22 01:20:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

