

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082124\
 Data File : VY019198.D
 Acq On : 22 Aug 2024 04:33
 Operator : SY/MD
 Sample : P3671-04
 Misc : 4.93g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-911-K1-SO-0.5-1-081924

Quant Time: Aug 22 06:29:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	82692	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.610	114	143091	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	117154	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	31892	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	42857	51.702	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.400%
35) Dibromofluoromethane	7.628	113	47750	51.036	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.080%
50) Toluene-d8	10.109	98	160644	47.712	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.420%
62) 4-Bromofluorobenzene	12.408	95	39364	34.296	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	68.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082124\
Data File : VY019198.D
Acq On : 22 Aug 2024 04:33
Operator : SY/MD
Sample : P3671-04
Misc : 4.93g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-911-K1-SO-0.5-1-081924

Quant Time: Aug 22 06:29:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

