

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082124\
 Data File : VY019208.D
 Acq On : 22 Aug 2024 08:32
 Operator : SY/MD
 Sample : P3671-14
 Misc : 6.18g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-912-J-SO-0.5-1-081924

Quant Time: Aug 22 08:53:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	19141	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.609	114	24717	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	16732	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.353	152	3275	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	6087	31.724	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	63.440%
35) Dibromofluoromethane	7.628	113	7944	49.154	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.300%
50) Toluene-d8	10.103	98	26590	45.719	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.440%
62) 4-Bromofluorobenzene	12.408	95	4900	24.714	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	49.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082124\
Data File : VY019208.D
Acq On : 22 Aug 2024 08:32
Operator : SY/MD
Sample : P3671-14
Misc : 6.18g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-912-J-SO-0.5-1-081924

Quant Time: Aug 22 08:53:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

