

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082124\
 Data File : VY019189.D
 Acq On : 21 Aug 2024 23:05
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Aug 22 01:22:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
 Supervised By :Semsettin Yesilyurt 08/23/2024

08/23/2024
 Supervised By :Semsettin
 Yesilyurt

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	117531	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	194042	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	172121	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	87346	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.055	65	64833	55.029	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 110.060%	
35) Dibromofluoromethane	7.628	113	64247	50.638	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.280%	
50) Toluene-d8	10.103	98	230767	50.542	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.080%	
62) 4-Bromofluorobenzene	12.407	95	78175	50.225	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 100.460%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.867	85	36978	42.883	ug/l	99
3) Chloromethane	2.062	50	72806	46.842	ug/l	95
4) Vinyl Chloride	2.202	62	82178	42.804	ug/l	98
5) Bromomethane	2.592	94	55087	46.590	ug/l	91
6) Chloroethane	2.732	64	51922	41.838	ug/l	95
7) Trichlorofluoromethane	3.049	101	117987	40.648	ug/l	98
8) Diethyl Ether	3.452	74	29628	47.289	ug/l	93
9) 1,1,2-Trichlorotrifluo...	3.811	101	55377	42.914	ug/l	99
10) Methyl Iodide	4.000	142	71902	56.847	ug/l	94
11) Tert butyl alcohol	4.860	59	20655	196.949	ug/l #	85
12) 1,1-Dichloroethene	3.781	96	54261	46.283	ug/l	93
13) Acrolein	3.647	56	16445	180.363	ug/l	92
14) Allyl chloride	4.378	41	78611	50.433	ug/l	91
15) Acrylonitrile	5.049	53	63562	247.942	ug/l	98
16) Acetone	3.860	43	54932	251.413	ug/l #	87
17) Carbon Disulfide	4.098	76	154291	46.660	ug/l	97
18) Methyl Acetate	4.378	43	33813	50.424	ug/l	94
19) Methyl tert-butyl Ether	5.110	73	145763	51.283	ug/l	97
20) Methylene Chloride	4.610	84	69605	51.629	ug/l	95
21) trans-1,2-Dichloroethene	5.104	96	64125	48.958	ug/l	95
22) Diisopropyl ether	6.012	45	188391	52.796	ug/l #	92
23) Vinyl Acetate	5.951	43	708837	250.910	ug/l	95
24) 1,1-Dichloroethane	5.909	63	109181	50.033	ug/l	100
25) 2-Butanone	6.884	43	86081	244.635	ug/l	93
26) 2,2-Dichloropropane	6.872	77	86710	49.069	ug/l	98
27) cis-1,2-Dichloroethene	6.884	96	77362	50.189	ug/l	97
28) Bromochloromethane	7.238	49	47167	51.527	ug/l	95
29) Tetrahydrofuran	7.256	42	54789	242.220	ug/l	90
30) Chloroform	7.408	83	125844	51.754	ug/l	99
31) Cyclohexane	7.689	56	81300	48.077	ug/l	98
32) 1,1,1-Trichloroethane	7.610	97	108599	50.820	ug/l	96
36) 1,1-Dichloropropene	7.829	75	82865	47.562	ug/l	99
37) Ethyl Acetate	6.976	43	39229	50.822	ug/l	97
38) Carbon Tetrachloride	7.811	117	97395	48.238	ug/l	95
39) Methylcyclohexane	9.103	83	88590	39.624	ug/l	97
40) Benzene	8.073	78	262284	49.086	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082124\
 Data File : VY019189.D
 Acq On : 21 Aug 2024 23:05
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Aug 22 01:22:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
 Supervised By :Semsettin Yesilyurt 08/23/2024

08/23/2024
 Supervised By :Semsettin
 Yesilyurt

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.213	41	20953	49.733	ug/l	91
42) 1,2-Dichloroethane	8.146	62	74196	51.421	ug/l	96
43) Isopropyl Acetate	8.189	43	76192	48.807	ug/l #	82
44) Trichloroethene	8.859	130	68446	45.908	ug/l	95
45) 1,2-Dichloropropane	9.140	63	52923	44.454	ug/l	94
46) Dibromomethane	9.225	93	34435	44.596	ug/l	99
47) Bromodichloromethane	9.420	83	93898	49.566	ug/l	99
48) Methyl methacrylate	9.219	41	30517	43.864	ug/l	90
49) 1,4-Dioxane	9.225	88	7753	828.872	ug/l #	97
51) 4-Methyl-2-Pentanone	9.999	43	204630	246.972	ug/l	92
52) Toluene	10.170	92	175982	50.173	ug/l	94
53) t-1,3-Dichloropropene	10.396	75	81163	51.648	ug/l	98
54) cis-1,3-Dichloropropene	9.853	75	94299	49.896	ug/l #	88
55) 1,1,2-Trichloroethane	10.572	97	48776	49.069	ug/l	97
56) Ethyl methacrylate	10.438	69	65071	50.012	ug/l #	82
57) 1,3-Dichloropropane	10.719	76	80547	49.707	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.707	63	151093	244.789	ug/l	97
59) 2-Hexanone	10.761	43	141106	247.888	ug/l	90
60) Dibromochloromethane	10.914	129	67875	50.133	ug/l	100
61) 1,2-Dibromoethane	11.017	107	48803	50.796	ug/l	95
64) Tetrachloroethene	10.646	164	75642	50.278	ug/l	98
65) Chlorobenzene	11.444	112	191565	48.335	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.517	131	69393	49.788	ug/l	96
67) Ethyl Benzene	11.517	91	331004	48.118	ug/l	97
68) m/p-Xylenes	11.627	106	263931	96.696	ug/l	99
69) o-Xylene	11.956	106	128293	50.756	ug/l	96
70) Styrene	11.969	104	216352	50.194	ug/l	97
71) Bromoform	12.133	173	40638	51.006	ug/l #	97
73) Isopropylbenzene	12.255	105	321213	47.981	ug/l	100
74) N-amyl acetate	12.072	43	72003	49.323	ug/l	91
75) 1,1,2,2-Tetrachloroethane	12.511	83	53939	45.257	ug/l	99
76) 1,2,3-Trichloropropane	12.560	75	36344m	42.192	ug/l	
77) Bromobenzene	12.536	156	76439	47.543	ug/l	99
78) n-propylbenzene	12.596	91	379735	47.449	ug/l	99
79) 2-Chlorotoluene	12.682	91	217621	48.502	ug/l	100
80) 1,3,5-Trimethylbenzene	12.737	105	265453	48.673	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.304	75	16588	47.559	ug/l #	84
82) 4-Chlorotoluene	12.779	91	226153	48.447	ug/l	99
83) tert-Butylbenzene	12.999	119	235243	46.708	ug/l	97
84) 1,2,4-Trimethylbenzene	13.048	105	265563	48.856	ug/l	99
85) sec-Butylbenzene	13.176	105	343332	46.779	ug/l	98
86) p-Isopropyltoluene	13.298	119	292257	48.478	ug/l	97
87) 1,3-Dichlorobenzene	13.291	146	150213	47.375	ug/l	99
88) 1,4-Dichlorobenzene	13.371	146	149893	47.557	ug/l	98
89) n-Butylbenzene	13.621	91	255527	46.490	ug/l	99
90) Hexachloroethane	13.883	117	56210	47.304	ug/l	100
91) 1,2-Dichlorobenzene	13.663	146	132115	47.376	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.279	75	8563	48.839	ug/l	91
93) 1,2,4-Trichlorobenzene	14.925	180	73259	46.583	ug/l	97
94) Hexachlorobutadiene	15.029	225	40731	49.724	ug/l	98
95) Naphthalene	15.151	128	131970	47.020	ug/l	100
96) 1,2,3-Trichlorobenzene	15.334	180	63771	48.078	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082124\
Data File : VY019189.D
Acq On : 21 Aug 2024 23:05
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Quant Time: Aug 22 01:22:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024

08/23/2024
Supervised By :Semsettin
Yesilyurt

08/23/2024


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.701 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:168 Resp: 11753
Ion Ratio Lower Upper
168 100
99 54.2 39.1 58.7

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#2
Dichlorodifluoromethane
Concen: 42.883 ug/l
RT: 1.867 min Scan# 24
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 85 Resp: 36978
Ion Ratio Lower Upper
85 100
87 31.5 16.0 48.0





#3
Chloromethane
Concen: 46.842 ug/l
RT: 2.062 min Scan# 50
Delta R.T. -0.000 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 50 Resp: 72800
Ion Ratio Lower Upper
50 100
52 34.7 25.6 38.4

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#4
Vinyl Chloride
Concen: 42.804 ug/l
RT: 2.202 min Scan# 79
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 62 Resp: 82178
Ion Ratio Lower Upper
62 100
64 31.6 26.1 39.1





#5
Bromomethane
Concen: 46.590 ug/l
RT: 2.592 min Scan# 142
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC





Tgt Ion: 94 Resp: 5508
Ion Ratio Lower Upper
94 100
96 101.6 74.1 111.1

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#6
Chloroethane
Concen: 41.838 ug/l
RT: 2.732 min Scan# 166
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 64 Resp: 51922
Ion Ratio Lower Upper
64 100
66 35.3 26.1 39.1





#7
 Trichlorofluoromethane
 Concen: 40.648 ug/l
 RT: 3.049 min Scan# 217
 Delta R.T. 0.006 min
 Lab File: VY019189.D
 Acq: 21 Aug 2024 23:05

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC





Tgt Ion:101 Resp: 11798
 Ion Ratio Lower Upper
 101 100
 103 63.4 49.3 73.9

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
 Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
 Supervised By :Semsettin
 Yesilyurt


08/23/2024



#8
 Diethyl Ether
 Concen: 47.289 ug/l
 RT: 3.452 min Scan# 284
 Delta R.T. 0.012 min
 Lab File: VY019189.D
 Acq: 21 Aug 2024 23:05

Tgt Ion: 74 Resp: 29628
 Ion Ratio Lower Upper
 74 100
 45 89.1 41.3 124.1





#9
1,1,2-Trichlorotrifluoroethane
Concen: 42.914 ug/l
RT: 3.811 min Scan# 341
Delta R.T. 0.012 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC





Tgt Ion:101 Resp: 5537
Ion Ratio Lower Upper
101 100
85 44.9 34.5 51.7
151 77.4 61.7 92.5

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#10
Methyl Iodide
Concen: 56.847 ug/l
RT: 4.000 min Scan# 374
Delta R.T. 0.012 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Tgt Ion:142 Resp: 71902
Ion Ratio Lower Upper
142 100
127 46.4 32.8 49.2
141 14.3 11.4 17.2







#11
Tert butyl alcohol
Concen: 196.949 ug/l
RT: 4.860 min Scan# 512
Delta R.T. 0.018 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 59 Resp: 2065
Ion Ratio Lower Upper
59 100
57 8.6 3.0 4.6



Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024






08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#12
1,1-Dichloroethene
Concen: 46.283 ug/l
RT: 3.781 min Scan# 338
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 96 Resp: 54261
Ion Ratio Lower Upper
96 100
61 152.2 114.6 171.8
98 59.7 50.8 76.2





#13
Acrolein
Concen: 180.363 ug/l
RT: 3.647 min Scan# 314
Delta R.T. 0.012 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 56 Resp: 16445
Ion Ratio Lower Upper
56 100
55 76.2 55.7 83.5


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt



08/23/2024




#14
Allyl chloride
Concen: 50.433 ug/l
RT: 4.378 min Scan# 436
Delta R.T. 0.012 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05




Tgt Ion: 41 Resp: 78611
Ion Ratio Lower Upper
41 100
39 79.7 58.3 87.5
76 41.4 39.0 58.6





#15
Acrylonitrile
Concen: 247.942 ug/l
RT: 5.049 min Scan# 54
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 53 Resp: 6356
Ion Ratio Lower Upper
53 100
52 80.7 64.7 97.1
51 38.5 28.4 42.6

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#16
Acetone
Concen: 251.413 ug/l
RT: 3.860 min Scan# 351
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 43 Resp: 54932
Ion Ratio Lower Upper
43 100
58 31.4 31.7 47.5#





#17
Carbon Disulfide
Concen: 46.660 ug/l
RT: 4.098 min Scan# 390
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 76 Resp: 15429
Ion Ratio Lower Upper
76 100
78 9.6 6.8 10.2


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024






08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#18
Methyl Acetate
Concen: 50.424 ug/l
RT: 4.378 min Scan# 436
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Tgt Ion: 43 Resp: 33813
Ion Ratio Lower Upper
43 100
74 25.7 23.3 34.9





#19
Methyl tert-butyl Ether
Concen: 51.283 ug/l
RT: 5.110 min Scan# 5110
Delta R.T. 0.012 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC





Tgt Ion: 73 Resp: 14576
Ion Ratio Lower Upper
73 100
57 20.1 17.3 25.9

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#20
Methylene Chloride
Concen: 51.629 ug/l
RT: 4.610 min Scan# 474
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 84 Resp: 69605
Ion Ratio Lower Upper
84 100
49 109.5 84.4 126.6
51 37.3 26.0 39.0
86 65.6 49.5 74.3





#21
trans-1,2-Dichloroethene
Concen: 48.958 ug/l
RT: 5.104 min Scan# 511
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC





Tgt Ion: 96 Resp: 64125
Ion Ratio Lower Upper
96 100
61 129.1 97.5 146.3
98 62.2 50.5 75.7

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#22
Diisopropyl ether
Concen: 52.796 ug/l
RT: 6.012 min Scan# 704
Delta R.T. 0.012 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Tgt Ion: 45 Resp: 188391
Ion Ratio Lower Upper
45 100
43 61.1 42.2 63.2
87 31.6 25.0 37.6
59 10.3 10.5 15.7#







#23
 Vinyl Acetate
 Concen: 250.910 ug/l
 RT: 5.951 min Scan# 694
 Delta R.T. 0.006 min
 Lab File: VY019189.D
 Acq: 21 Aug 2024 23:05

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Tgt Ion: 43 Resp: 70883
 Ion Ratio Lower Upper
 43 100
 86 11.5 10.6 16.0



Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
 Supervised By :Semsettin Yesilyurt 08/23/2024






08/23/2024
 Supervised By :Semsettin
 Yesilyurt


08/23/2024



#24
 1,1-Dichloroethane
 Concen: 50.033 ug/l
 RT: 5.909 min Scan# 687
 Delta R.T. 0.006 min
 Lab File: VY019189.D
 Acq: 21 Aug 2024 23:05

Tgt Ion: 63 Resp: 109181
 Ion Ratio Lower Upper
 63 100
 98 8.1 4.0 12.2
 100 5.2 2.6 8.0







#25
2-Butanone
Concen: 244.635 ug/l
RT: 6.884 min Scan# 845
Delta R.T. 0.012 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 8608
Ion Ratio Lower Upper
43 100
72 30.1 27.3 40.9



Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024






08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#26
2,2-Dichloropropane
Concen: 49.069 ug/l
RT: 6.872 min Scan# 845
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 77 Resp: 86710
Ion Ratio Lower Upper
77 100
97 23.3 12.0 36.1





#27
 cis-1,2-Dichloroethene
 Concen: 50.189 ug/l
 RT: 6.884 min Scan# 845
 Delta R.T. 0.012 min
 Lab File: VY019189.D
 Acq: 21 Aug 2024 23:05

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC






Tgt Ion: 96 Resp: 77362
 Ion Ratio Lower Upper
 96 100
 61 134.1 0.0 256.2
 98 65.1 0.0 130.0

Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
 Supervised By :Semsettin Yesilyurt 08/23/2024

08/23/2024
 Supervised By :Semsettin
 Yesilyurt



08/23/2024




#28
 Bromochloromethane
 Concen: 51.527 ug/l
 RT: 7.238 min Scan# 905
 Delta R.T. 0.012 min
 Lab File: VY019189.D
 Acq: 21 Aug 2024 23:05



Tgt Ion: 49 Resp: 47167
 Ion Ratio Lower Upper
 49 100
 129 2.3 0.0 4.6
 130 86.5 72.7 109.1





#29

Tetrahydrofuran

Concen: 242.220 ug/l

RT: 7.256 min Scan# 906

Delta R.T. 0.012 min

Lab File: VY019189.D

Acq: 21 Aug 2024 23:05

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 42	Resp: 54789
Ion Ratio	Lower Upper
42 100	
72 46.3	43.0 64.4
71 44.2	40.8 61.2

Manual Integrations





APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024

08/23/2024

Supervised By :Semsettin Yesilyurt

08/23/2024



#30

Chloroform

Concen: 51.754 ug/l




RT: 7.408 min Scan# 933


Delta R.T. 0.000 min

Lab File: VY019189.D

Acq: 21 Aug 2024 23:05


Tgt Ion: 83	Resp: 125844
Ion Ratio	Lower Upper
83 100	
85 64.6	52.2 78.4





#31
Cyclohexane
Concen: 48.077 ug/l
RT: 7.689 min Scan# 911
Delta R.T. -0.000 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 56 Resp: 81300
Ion Ratio Lower Upper
56 100
69 36.1 28.7 43.1
84 94.5 77.7 116.5

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#32
1,1,1-Trichloroethane
Concen: 50.820 ug/l
RT: 7.610 min Scan# 966
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Tgt Ion: 97 Resp: 108599
Ion Ratio Lower Upper
97 100
99 65.7 50.3 75.5
61 43.1 32.6 48.8





#33
1,2-Dichloroethane-d4
Concen: 55.029 ug/l
RT: 8.055 min Scan# 1130
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 6483
Ion Ratio Lower Upper
65 100
67 51.7 0.0 109.6

Manual Integrations





APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024

08/23/2024




Supervised By :Semsettin
Yesilyurt


08/23/2024



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.609 min Scan# 1130
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion:114 Resp: 194042
Ion Ratio Lower Upper
114 100
63 18.9 0.0 35.0
88 14.0 0.0 27.2





#35
Dibromofluoromethane
Concen: 50.638 ug/l
RT: 7.628 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC





Tgt Ion:113 Resp: 6424
Ion Ratio Lower Upper
113 100
111 101.7 82.2 123.4
192 20.2 15.9 23.9

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#36
1,1-Dichloropropene
Concen: 47.562 ug/l
RT: 7.829 min Scan# 1002
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 75 Resp: 82865
Ion Ratio Lower Upper
75 100
110 37.3 19.0 57.0
77 31.2 25.3 37.9





#37
Ethyl Acetate
Concen: 50.822 ug/l
RT: 6.976 min Scan# 80
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC





Tgt Ion: 43 Resp: 39229
Ion Ratio Lower Upper
43 100
61 14.3 12.6 19.0
70 10.3 7.8 11.8

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#38
Carbon Tetrachloride
Concen: 48.238 ug/l
RT: 7.811 min Scan# 999
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion:117 Resp: 97395
Ion Ratio Lower Upper
117 100
119 100.5 75.8 113.6
121 28.7 24.0 36.0





#39
Methylcyclohexane
Concen: 39.624 ug/l
RT: 9.103 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 83 Resp: 88590
Ion Ratio Lower Upper
83 100
55 67.8 51.5 77.3
98 51.8 40.5 60.7

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#40
Benzene
Concen: 49.086 ug/l
RT: 8.073 min Scan# 1042
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 78 Resp: 262284
Ion Ratio Lower Upper
78 100
77 22.9 19.5 29.3





#41
Methacrylonitrile
Concen: 49.733 ug/l
RT: 7.213 min Scan# 904
Delta R.T. -0.018 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 41 Resp: 2095
Ion Ratio Lower Upper
41 100
39 55.8 48.0 72.0
67 80.5 75.1 112.7
52 34.6 29.8 44.6

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#42
1,2-Dichloroethane
Concen: 51.421 ug/l
RT: 8.146 min Scan# 1054
Delta R.T. 0.000 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Tgt Ion: 62 Resp: 74196
Ion Ratio Lower Upper
62 100
98 9.5 0.0 22.0





#43

Isopropyl Acetate

Concen: 48.807 ug/l

RT: 8.189 min Scan# 1060

Delta R.T. 0.006 min

Lab File: VY019189.D

Acq: 21 Aug 2024 23:05

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43	Resp: 76192
Ion Ratio	Lower Upper
43 100	
61 20.8	27.9 41.9
87 14.3	13.1 19.7

Manual Integrations





APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024

08/23/2024

Supervised By :Semsettin Yesilyurt

08/23/2024



#44

Trichloroethene

Concen: 45.908 ug/l




RT: 8.859 min Scan# 1171


Delta R.T. 0.006 min

Lab File: VY019189.D

Acq: 21 Aug 2024 23:05

Tgt Ion:130	Resp: 68446
Ion Ratio	Lower Upper
130 100	
95 98.6	0.0 188.0





#45

1,2-Dichloropropane

Concen: 44.454 ug/l

RT: 9.140 min Scan# 1217

Delta R.T. 0.012 min

Lab File: VY019189.D

Acq: 21 Aug 2024 23:05

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 5292

Ion Ratio Lower Upper

63 100

65 28.6 25.4 38.2

Manual Integrations

APPROVED





Reviewed By :Mahesh Dadoda 08/23/2024

Supervised By :Semsettin Yesilyurt 08/23/2024

08/23/2024

Supervised By :Semsettin
Yesilyurt

08/23/2024



#46

Dibromomethane

Concen: 44.596 ug/l

RT: 9.225 min Scan# 1231

Delta R.T. 0.006 min

Lab File: VY019189.D

Acq: 21 Aug 2024 23:05




Tgt Ion: 93 Resp: 34435


Ion Ratio Lower Upper

93 100

95 83.6 66.2 99.4

174 100.2 80.9 121.3





#47
Bromodichloromethane
Concen: 49.566 ug/l
RT: 9.420 min Scan# 1262
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC





Tgt Ion: 83 Resp: 9389
Ion Ratio Lower Upper
83 100
85 62.0 50.1 75.1
127 9.0 7.4 11.0

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#48
Methyl methacrylate
Concen: 43.864 ug/l
RT: 9.219 min Scan# 1230
Delta R.T. 0.012 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 41 Resp: 30517
Ion Ratio Lower Upper
41 100
69 97.0 84.1 126.1
39 64.7 43.6 65.4





#49
1,4-Dioxane
Concen: 828.872 ug/l
RT: 9.225 min Scan# 1229
Delta R.T. 0.012 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 88 Resp: 775
Ion Ratio Lower Upper
88 100
43 32.0 0.0 0.0
58 59.7 49.3 73.9

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#50
Toluene-d8
Concen: 50.542 ug/l
RT: 10.103 min Scan# 1375
Delta R.T. 0.000 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Tgt Ion: 98 Resp: 230767
Ion Ratio Lower Upper
98 100
100 67.0 52.0 78.0







#51
4-Methyl-2-Pentanone
Concen: 246.972 ug/l
RT: 9.999 min Scan# 1357
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 204630
Ion Ratio Lower Upper
43 100
58 39.6 35.7 53.5



Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024






08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#52
Toluene
Concen: 50.173 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 92 Resp: 175982
Ion Ratio Lower Upper
92 100
91 165.8 139.0 208.6





#53
t-1,3-Dichloropropene
Concen: 51.648 ug/l
RT: 10.396 min Scan# 1421
Delta R.T. 0.012 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 75 Resp: 8116
Ion Ratio Lower Upper
75 100
77 31.4 24.1 36.1

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#54
cis-1,3-Dichloropropene
Concen: 49.896 ug/l
RT: 9.853 min Scan# 1334
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 75 Resp: 94299
Ion Ratio Lower Upper
75 100
77 33.3 25.4 38.0
39 49.3 29.8 44.6#





#55
1,1,2-Trichloroethane
Concen: 49.069 ug/l
RT: 10.572 min Scan# 1451
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 97 Resp: 48770
Ion Ratio Lower Upper
97 100
83 82.4 67.5 101.3
85 51.4 42.9 64.3
99 60.9 46.6 70.0

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#56
Ethyl methacrylate
Concen: 50.012 ug/l
RT: 10.438 min Scan# 1430
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 69 Resp: 65071
Ion Ratio Lower Upper
69 100
41 63.8 42.3 63.5#
39 40.2 22.9 34.3#





#57
1,3-Dichloropropane
Concen: 49.707 ug/l
RT: 10.719 min Scan# 1475
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 76 Resp: 8054
Ion Ratio Lower Upper
76 100
78 32.7 26.2 39.4


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024






08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#58
2-Chloroethyl Vinyl ether
Concen: 244.789 ug/l
RT: 9.707 min Scan# 1310
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 63 Resp: 151093
Ion Ratio Lower Upper
63 100
106 31.1 23.7 35.5





#59
2-Hexanone
Concen: 247.888 ug/l
RT: 10.761 min Scan# 1482
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 43 Resp: 141100
Ion Ratio Lower Upper
43 100
58 54.7 31.4 94.0

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#60
Dibromochloromethane
Concen: 50.133 ug/l
RT: 10.914 min Scan# 1508
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion:129 Resp: 67875
Ion Ratio Lower Upper
129 100
127 77.9 39.0 117.0





#61
1,2-Dibromoethane
Concen: 50.796 ug/l
RT: 11.017 min Scan# 1107
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05



Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion:107 Resp: 4880
Ion Ratio Lower Upper
107 100
109 90.2 75.9 113.9


Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024






08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#62
4-Bromofluorobenzene
Concen: 50.225 ug/l
RT: 12.407 min Scan# 1753
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 95 Resp: 78175
Ion Ratio Lower Upper
95 100
174 84.9 0.0 175.6
176 86.7 0.0 171.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.420 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:117 Resp: 17212
Ion Ratio Lower Upper
117 100
82 50.5 42.4 63.6
119 31.6 25.9 38.9

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#64
Tetrachloroethene
Concen: 50.278 ug/l
RT: 10.646 min Scan# 1464
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion:164 Resp: 75642
Ion Ratio Lower Upper
164 100
166 124.0 102.6 154.0
129 88.7 72.0 108.0
131 88.8 70.0 105.0





#65
Chlorobenzene
Concen: 48.335 ug/l
RT: 11.444 min Scan# 111
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:112 Resp: 19156
Ion Ratio Lower Upper
112 100
114 32.3 26.4 39.6

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 49.788 ug/l
RT: 11.517 min Scan# 1607
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Tgt Ion:131 Resp: 69393
Ion Ratio Lower Upper
131 100
133 93.7 47.8 143.3
119 62.0 34.1 102.2







#67
Ethyl Benzene
Concen: 48.118 ug/l
RT: 11.517 min Scan# 1607
Delta R.T. 0.000 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 331004
Ion Ratio Lower Upper
91 100
106 32.1 24.5 36.7



Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024






08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#68
m/p-Xylenes
Concen: 96.696 ug/l
RT: 11.627 min Scan# 1625
Delta R.T. 0.000 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Tgt Ion:106 Resp: 263931
Ion Ratio Lower Upper
106 100
91 196.0 156.0 234.0







#69
o-Xylene
Concen: 50.756 ug/l
RT: 11.956 min Scan# 1678
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:106 Resp: 12829
Ion Ratio Lower Upper
106 100
91 201.3 103.6 310.8



Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024






08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#70
Styrene
Concen: 50.194 ug/l
RT: 11.969 min Scan# 1681
Delta R.T. 0.000 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Tgt Ion:104 Resp: 216352
Ion Ratio Lower Upper
104 100
78 48.5 37.0 55.4
103 53.7 44.2 66.2





#71

Bromoform

Concen: 51.006 ug/l

RT: 12.133 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY019189.D

Acq: 21 Aug 2024 23:05

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion:173	Resp: 40638
Ion Ratio	Lower Upper
173	100
175	47.8 25.1 75.3
254	0.0 0.0 0.0

Manual Integrations





APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024

08/23/2024

Supervised By :Semsettin Yesilyurt

08/23/2024



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.352 min Scan# 1908


Delta R.T. 0.006 min

Lab File: VY019189.D

Acq: 21 Aug 2024 23:05

Tgt Ion:152	Resp: 87346
Ion Ratio	Lower Upper
152	100
115	57.4 28.2 84.7
150	170.7 0.0 345.6





#73
Isopropylbenzene
Concen: 47.981 ug/l
RT: 12.255 min Scan# 1728
Delta R.T. 0.000 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion:105 Resp: 32121
Ion Ratio Lower Upper
105 100
120 26.6 13.4 40.1

Manual Integrations





APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024

08/23/2024




Supervised By :Semsettin
Yesilyurt


08/23/2024



#74
N-ethyl acetate
Concen: 49.323 ug/l
RT: 12.072 min Scan# 1698
Delta R.T. 0.000 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 43 Resp: 72003
Ion Ratio Lower Upper
43 100
70 43.2 41.4 62.2
55 25.0 23.3 34.9
61 24.9 22.6 33.8





#75
1,1,2,2-Tetrachloroethane
Concen: 45.257 ug/l
RT: 12.511 min Scan# 1770
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 83 Resp: 53939
Ion Ratio Lower Upper
83 100
131 11.1 5.0 14.9
85 63.4 32.0 96.0

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#76
1,2,3-Trichloropropane
Concen: 42.192 ug/l m
RT: 12.560 min Scan# 1778
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 75 Resp: 36344
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77
Bromobenzene
Concen: 47.543 ug/l
RT: 12.536 min Scan# 1774
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:156 Resp: 76439
Ion Ratio Lower Upper
156 100
77 171.7 87.1 261.3
158 98.0 48.6 145.9

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024

08/23/2024
Supervised By :Semsettin
Yesilyurt



08/23/2024




#78
n-propylbenzene
Concen: 47.449 ug/l
RT: 12.596 min Scan# 1784
Delta R.T. 0.000 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05




Tgt Ion: 91 Resp: 379735
Ion Ratio Lower Upper
91 100
120 23.4 11.5 34.4





#79
2-Chlorotoluene
Concen: 48.502 ug/l
RT: 12.682 min Scan# 1798
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion: 91 Resp: 21762
Ion Ratio Lower Upper
91 100
126 35.5 17.8 53.5

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#80
1,3,5-Trimethylbenzene
Concen: 48.673 ug/l
RT: 12.737 min Scan# 1807
Delta R.T. 0.000 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion:105 Resp: 265453
Ion Ratio Lower Upper
105 100
120 50.5 25.1 75.2





#81
trans-1,4-Dichloro-2-butene
Concen: 47.559 ug/l
RT: 12.304 min Scan# 1736
Delta R.T. 0.000 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC





Tgt Ion: 75 Resp: 1658
Ion Ratio Lower Upper
75 100
53 104.7 66.8 100.2
89 46.0 35.8 53.6

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#82
4-Chlorotoluene
Concen: 48.447 ug/l
RT: 12.779 min Scan# 1814
Delta R.T. 0.000 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 91 Resp: 226153
Ion Ratio Lower Upper
91 100
126 35.7 17.6 52.9





#83
tert-Butylbenzene
Concen: 46.708 ug/l
RT: 12.999 min Scan# 1850
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC





Tgt Ion:119 Resp: 23524
Ion Ratio Lower Upper
119 100
91 65.9 31.3 93.8
134 27.1 13.4 40.1

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#84
1,2,4-Trimethylbenzene
Concen: 48.856 ug/l
RT: 13.048 min Scan# 1858
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion:105 Resp: 265563
Ion Ratio Lower Upper
105 100
120 46.8 23.2 69.6





#85
 sec-Butylbenzene
 Concen: 46.779 ug/l
 RT: 13.176 min Scan# 1879
 Delta R.T. 0.000 min
 Lab File: VY019189.D
 Acq: 21 Aug 2024 23:05

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC






Tgt Ion:105 Resp: 34333
 Ion Ratio Lower Upper
 105 100
 134 21.1 10.1 30.1

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
 Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
 Supervised By :Semsettin
 Yesilyurt


08/23/2024



#86
 p-Isopropyltoluene
 Concen: 48.478 ug/l
 RT: 13.298 min Scan# 1899
 Delta R.T. 0.006 min
 Lab File: VY019189.D
 Acq: 21 Aug 2024 23:05


Tgt Ion:119 Resp: 292257
 Ion Ratio Lower Upper
 119 100
 134 27.3 13.1 39.3
 91 23.5 10.7 32.1





#87
1,3-Dichlorobenzene
Concen: 47.375 ug/l
RT: 13.291 min Scan# 1897
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:146 Resp: 15021
Ion Ratio Lower Upper
146 100
111 39.1 19.0 57.0
148 64.0 31.7 95.1

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#88
1,4-Dichlorobenzene
Concen: 47.557 ug/l
RT: 13.371 min Scan# 1911
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Tgt Ion:146 Resp: 149893
Ion Ratio Lower Upper
146 100
111 37.3 18.4 55.4
148 62.8 32.4 97.2





#89
n-Butylbenzene
Concen: 46.490 ug/l
RT: 13.621 min Scan# 1952
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC





Tgt Ion: 91 Resp: 25552
Ion Ratio Lower Upper
91 100
92 53.1 26.6 79.7
134 26.9 12.9 38.7

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#90
Hexachloroethane
Concen: 47.304 ug/l
RT: 13.883 min Scan# 1995
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion:117 Resp: 56210
Ion Ratio Lower Upper
117 100
201 74.5 37.4 112.2





#91
1,2-Dichlorobenzene
Concen: 47.376 ug/l
RT: 13.663 min Scan# 1958
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:146 Resp: 13211
Ion Ratio Lower Upper
146 100
111 40.1 19.9 59.7
148 64.0 31.8 95.4

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 48.839 ug/l
RT: 14.279 min Scan# 2060
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion: 75 Resp: 8563
Ion Ratio Lower Upper
75 100
155 90.9 48.6 145.8
157 112.8 62.7 188.2





#93
1,2,4-Trichlorobenzene
Concen: 46.583 ug/l
RT: 14.925 min Scan# 2166
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:180 Resp: 73259
Ion Ratio Lower Upper
180 100
182 93.5 47.9 143.8
145 29.9 14.2 42.6

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024




08/23/2024
Supervised By :Semsettin
Yesilyurt


08/23/2024



#94
Hexachlorobutadiene
Concen: 49.724 ug/l
RT: 15.029 min Scan# 2183
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05


Tgt Ion:225 Resp: 40731
Ion Ratio Lower Upper
225 100
223 63.2 30.4 91.2
227 64.2 31.9 95.9





#95
Naphthalene
Concen: 47.020 ug/l
RT: 15.151 min Scan# 2203
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC






Tgt Ion:128 Resp: 131970
Ion Ratio Lower Upper
128 100
127 13.0 10.6 16.0
129 10.7 8.6 13.0

Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 08/23/2024
Supervised By :Semsettin Yesilyurt 08/23/2024

08/23/2024
Supervised By :Semsettin
Yesilyurt

08/23/2024



#96
1,2,3-Trichlorobenzene
Concen: 48.078 ug/l
RT: 15.334 min Scan# 2233
Delta R.T. 0.006 min
Lab File: VY019189.D
Acq: 21 Aug 2024 23:05



Tgt Ion:180 Resp: 63771
Ion Ratio Lower Upper
180 100
182 94.9 47.6 142.9
145 31.9 15.4 46.1

