

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082525\
 Data File : VY023208.D
 Acq On : 25 Aug 2025 12:59
 Operator : SY/MD
 Sample : Q2915-03
 Misc : 5.51g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BRIDGE-EAST

Quant Time: Aug 25 23:50:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 13 02:10:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	7162	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.615	114	9208	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	3801	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.340	152	734	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	2929	42.910	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	85.820%
35) Dibromofluoromethane	7.628	113	2981	54.104	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.200%
50) Toluene-d8	10.109	98	8644	40.159	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	80.320%
62) 4-Bromofluorobenzene	12.401	95	808	11.673	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	23.340%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082525\
Data File : VY023208.D
Acq On : 25 Aug 2025 12:59
Operator : SY/MD
Sample : Q2915-03
Misc : 5.51g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BRIDGE-EAST

Quant Time: Aug 25 23:50:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 13 02:10:11 2025
Response via : Initial Calibration

