

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082824\
 Data File : VY019333.D
 Acq On : 28 Aug 2024 16:26
 Operator : SY/MD
 Sample : P3741-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-05-08262024

Quant Time: Aug 29 01:35:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	14162	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	15747	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	17464	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	5098	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	5991	42.201	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	84.400%
35) Dibromofluoromethane	7.628	113	6730	65.363	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	130.720%
50) Toluene-d8	10.103	98	21537	58.125	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	116.240%
62) 4-Bromofluorobenzene	12.408	95	6037	47.794	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	95.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082824\
Data File : VY019333.D
Acq On : 28 Aug 2024 16:26
Operator : SY/MD
Sample : P3741-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-05-08262024

Quant Time: Aug 29 01:35:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

