

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082824\
 Data File : VY019334.D
 Acq On : 28 Aug 2024 16:52
 Operator : SY/MD
 Sample : P3749-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CL-01-082624

Quant Time: Aug 29 01:36:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	127602	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	228031	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	208045	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	83225	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	71266	55.715	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.440%
35) Dibromofluoromethane	7.622	113	70774	47.467	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.940%
50) Toluene-d8	10.103	98	254145	47.366	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.740%
62) 4-Bromofluorobenzene	12.408	95	80349	43.928	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	87.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082824\
Data File : VY019334.D
Acq On : 28 Aug 2024 16:52
Operator : SY/MD
Sample : P3749-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CL-01-082624

Quant Time: Aug 29 01:36:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

