

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082824\
 Data File : VY019335.D
 Acq On : 28 Aug 2024 17:17
 Operator : SY/MD
 Sample : P3748-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-02-082624

Quant Time: Aug 29 01:36:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	60138	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	106723	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	96016	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	35261	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	35108	58.238	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.480%
35) Dibromofluoromethane	7.622	113	35597	51.012	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.020%
50) Toluene-d8	10.103	98	120244	47.883	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.760%
62) 4-Bromofluorobenzene	12.408	95	34461	40.255	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	80.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082824\
Data File : VY019335.D
Acq On : 28 Aug 2024 17:17
Operator : SY/MD
Sample : P3748-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-02-082624

Quant Time: Aug 29 01:36:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

