

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082824\
 Data File : VY019338.D
 Acq On : 28 Aug 2024 18:32
 Operator : SY/MD
 Sample : P3767-02
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 P-1-COMPOSITE

Quant Time: Aug 29 01:37:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	15865	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.604	114	24955	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	24485	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	6585	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.049	65	6113	38.438	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	76.880%
35) Dibromofluoromethane	7.622	113	8672	53.147	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.300%
50) Toluene-d8	10.103	98	34028	57.950	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	115.900%
62) 4-Bromofluorobenzene	12.408	95	7796	38.946	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	77.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082824\
Data File : VY019338.D
Acq On : 28 Aug 2024 18:32
Operator : SY/MD
Sample : P3767-02
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
P-1-COMPOSITE

Quant Time: Aug 29 01:37:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

