

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082824\
 Data File : VY019355.D
 Acq On : 29 Aug 2024 03:52
 Operator : SY/MD
 Sample : P3737-02
 Misc : 4.50g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-2

Quant Time: Aug 29 04:52:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	185581	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	328301	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	301442	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	113240	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	102836	55.279	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.560%
35) Dibromofluoromethane	7.628	113	104482	48.673	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.340%
50) Toluene-d8	10.103	98	373849	48.395	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.780%
62) 4-Bromofluorobenzene	12.408	95	119584	45.410	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	90.820%

Target Compounds			Qvalue			
16) Acetone	3.866	43	1608	4.661	ug/l	# 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082824\
Data File : VY019355.D
Acq On : 29 Aug 2024 03:52
Operator : SY/MD
Sample : P3737-02
Misc : 4.50g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-2

Quant Time: Aug 29 04:52:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

