

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082824\
 Data File : VY019356.D
 Acq On : 29 Aug 2024 04:16
 Operator : SY/MD
 Sample : P3737-03
 Misc : 5.15g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-3

Quant Time: Aug 29 04:53:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 05:10:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	186829	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	331640	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	301339	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	115204	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	100985	53.922	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.840%
35) Dibromofluoromethane	7.628	113	104341	48.118	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.240%
50) Toluene-d8	10.103	98	380347	48.740	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.480%
62) 4-Bromofluorobenzene	12.408	95	117342	44.110	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	88.220%
Target Compounds						
16) Acetone	3.860	43	2501	7.201	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082824\
Data File : VY019356.D
Acq On : 29 Aug 2024 04:16
Operator : SY/MD
Sample : P3737-03
Misc : 5.15g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-3

Quant Time: Aug 29 04:53:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 05:10:30 2024
Response via : Initial Calibration

