

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082925\
 Data File : VY023240.D
 Acq On : 29 Aug 2025 12:21
 Operator : SY/MD
 Sample : Q2971-25
 Misc : 5.44g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP7

Quant Time: Aug 30 01:20:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 13 02:10:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	558310	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	1077084	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	991867	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	396880	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	322798	60.664	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.320%
35) Dibromofluoromethane	7.634	113	357935	55.537	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.080%
50) Toluene-d8	10.109	98	1299252	51.603	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.200%
62) 4-Bromofluorobenzene	12.407	95	388563	47.988	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	95.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY082925\
Data File : VY023240.D
Acq On : 29 Aug 2025 12:21
Operator : SY/MD
Sample : Q2971-25
Misc : 5.44g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP7

Quant Time: Aug 30 01:20:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 13 02:10:11 2025
Response via : Initial Calibration

