

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083024\
 Data File : VY019404.D
 Acq On : 30 Aug 2024 18:33
 Operator : SY/MD
 Sample : P3741-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-05-08262024

Quant Time: Aug 30 23:18:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082924S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 02:45:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	102828	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	177520	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	152019	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	53235	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	50514	51.864	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.720%
35) Dibromofluoromethane	7.628	113	55153	49.141	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.280%
50) Toluene-d8	10.109	98	191246	46.289	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.580%
62) 4-Bromofluorobenzene	12.408	95	55820	41.772	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	83.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083024\
Data File : VY019404.D
Acq On : 30 Aug 2024 18:33
Operator : SY/MD
Sample : P3741-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-05-08262024

Quant Time: Aug 30 23:18:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082924S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 02:45:30 2024
Response via : Initial Calibration

