

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083024\
 Data File : VY019406.D
 Acq On : 30 Aug 2024 19:20
 Operator : SY/MD
 Sample : P3773-01
 Misc : 5.17g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 AU-06-082824

Quant Time: Aug 30 23:19:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082924S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 02:45:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	115654	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	212202	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	156195	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	46370	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	67607	61.716	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.440%
35) Dibromofluoromethane	7.634	113	68813	51.291	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.580%
50) Toluene-d8	10.109	98	241934	48.987	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.980%
62) 4-Bromofluorobenzene	12.408	95	57306	35.875	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	71.740%
Target Compounds						
16) Acetone	3.873	43	1742	8.085	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083024\
Data File : VY019406.D
Acq On : 30 Aug 2024 19:20
Operator : SY/MD
Sample : P3773-01
Misc : 5.17g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
AU-06-082824

Quant Time: Aug 30 23:19:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082924S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 02:45:30 2024
Response via : Initial Calibration

