

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083024\
 Data File : VY019413.D
 Acq On : 30 Aug 2024 22:04
 Operator : SY/MD
 Sample : P3737-02
 Misc : 5.27g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-2

Quant Time: Aug 30 23:22:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082924S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 02:45:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	111991	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	197984	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	178371	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	56653	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	59670	56.252	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.500%
35) Dibromofluoromethane	7.634	113	63943	51.084	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.160%
50) Toluene-d8	10.103	98	222106	48.202	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.400%
62) 4-Bromofluorobenzene	12.408	95	68309	45.834	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	91.660%
Target Compounds						
16) Acetone	3.867	43	1264m	6.058	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083024\
Data File : VY019413.D
Acq On : 30 Aug 2024 22:04
Operator : SY/MD
Sample : P3737-02
Misc : 5.27g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-2

Quant Time: Aug 30 23:22:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082924S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 02:45:30 2024
Response via : Initial Calibration

