

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083024\
 Data File : VY019415.D
 Acq On : 30 Aug 2024 22:51
 Operator : SY/MD
 Sample : P3737-05
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-5

Quant Time: Aug 31 01:17:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082924S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 02:45:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	89782	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	158390	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	113141	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	48980	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	46276	54.417	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.840%	
35) Dibromofluoromethane	7.634	113	51128	51.057	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.120%	
50) Toluene-d8	10.109	98	135553	36.772	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 73.540%	
62) 4-Bromofluorobenzene	12.408	95	47817	40.105	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 80.200%	
Target Compounds						
16) Acetone	3.873	43	2276	13.607	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY083024\
Data File : VY019415.D
Acq On : 30 Aug 2024 22:51
Operator : SY/MD
Sample : P3737-05
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-5

Quant Time: Aug 31 01:17:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082924S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 02:45:30 2024
Response via : Initial Calibration

