

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090225\
 Data File : VY023257.D
 Acq On : 02 Sep 2025 12:02
 Operator : SY/MD
 Sample : Q2982-03
 Misc : 5.37g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MOO-25-0244-0248

Quant Time: Sep 03 03:29:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 13 02:10:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	609376	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	1143088	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	1056918	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	430033	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	333454	57.415	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.840%
35) Dibromofluoromethane	7.634	113	372360	54.440	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.880%
50) Toluene-d8	10.109	98	1378739	51.598	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.200%
62) 4-Bromofluorobenzene	12.407	95	408857	47.579	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	95.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090225\
Data File : VY023257.D
Acq On : 02 Sep 2025 12:02
Operator : SY/MD
Sample : Q2982-03
Misc : 5.37g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MOO-25-0244-0248

Quant Time: Sep 03 03:29:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 13 02:10:11 2025
Response via : Initial Calibration

