

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090225\
 Data File : VY023265.D
 Acq On : 02 Sep 2025 15:10
 Operator : SY/MD
 Sample : Q2994-01
 Misc : 5.15g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PAULDING-WC1

Quant Time: Sep 03 03:32:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 13 02:10:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	590164	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1097764	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	927290	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	306386	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	306595	54.509	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.020%
35) Dibromofluoromethane	7.634	113	352346	53.640	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.280%
50) Toluene-d8	10.109	98	1265219	49.305	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.600%
62) 4-Bromofluorobenzene	12.402	95	335272	40.626	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090225\
Data File : VY023265.D
Acq On : 02 Sep 2025 15:10
Operator : SY/MD
Sample : Q2994-01
Misc : 5.15g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PAULDING-WC1

Quant Time: Sep 03 03:32:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 13 02:10:11 2025
Response via : Initial Calibration

