

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090425\
 Data File : VY023282.D
 Acq On : 04 Sep 2025 11:05
 Operator : SY/MD
 Sample : Q3009-01RE
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-04-090325RE

Quant Time: Sep 05 04:24:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 13 02:10:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	147128	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	235778	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	196732	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.347	152	71551	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	72494	51.699	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.400%
35) Dibromofluoromethane	7.634	113	84441	59.852	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	119.700%
50) Toluene-d8	10.109	98	272966	49.527	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.060%
62) 4-Bromofluorobenzene	12.408	95	68523	38.659	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090425\
Data File : VY023282.D
Acq On : 04 Sep 2025 11:05
Operator : SY/MD
Sample : Q3009-01RE
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-04-090325RE

Quant Time: Sep 05 04:24:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 13 02:10:11 2025
Response via : Initial Calibration

