

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090922\
 Data File : VY010438.D
 Acq On : 09 Sep 2022 23:13
 Operator : KP/MD
 Sample : N4509-05
 Misc : 5.49g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S1-SO_116D-20220902

Quant Time: Sep 12 03:02:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 09 17:08:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	200840	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	295800	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	216008	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	73620	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	53204	26.302	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	52.600%
35) Dibromofluoromethane	7.716	113	49518	23.571	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	47.140%#
50) Toluene-d8	10.179	98	335105	49.218	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.440%
62) 4-Bromofluorobenzene	12.483	95	76608	25.352	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	50.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090922\
Data File : VY010438.D
Acq On : 09 Sep 2022 23:13
Operator : KP/MD
Sample : N4509-05
Misc : 5.49g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S1-SO_116D-20220902

Quant Time: Sep 12 03:02:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 09 17:08:30 2022
Response via : Initial Calibration

