

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090924\
 Data File : VY019468.D
 Acq On : 09 Sep 2024 16:20
 Operator : SY/MD
 Sample : P3857-03RE
 Misc : 6.89g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-C1-GRABRE

Quant Time: Sep 10 01:49:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	266029	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	455477	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	360798	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	132582	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	124495	50.539	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.080%
35) Dibromofluoromethane	7.634	113	132619	51.354	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.700%
50) Toluene-d8	10.103	98	519893	54.428	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	108.860%
62) 4-Bromofluorobenzene	12.408	95	156320	42.561	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	85.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090924\
Data File : VY019468.D
Acq On : 09 Sep 2024 16:20
Operator : SY/MD
Sample : P3857-03RE
Misc : 6.89g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-C1-GRABRE

Quant Time: Sep 10 01:49:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

