

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090924\
 Data File : VY019470.D
 Acq On : 09 Sep 2024 17:07
 Operator : SY/MD
 Sample : P3905-05
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-1

Quant Time: Sep 10 01:49:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	482204	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.610	114	916544	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	819699	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	325339	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	266342	59.650	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 119.300%	
35) Dibromofluoromethane	7.634	113	285805	54.998	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.000%	
50) Toluene-d8	10.103	98	1084757	56.436	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 112.880%	
62) 4-Bromofluorobenzene	12.408	95	371428	50.256	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 100.520%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090924\
Data File : VY019470.D
Acq On : 09 Sep 2024 17:07
Operator : SY/MD
Sample : P3905-05
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-1

Quant Time: Sep 10 01:49:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

