

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091025\
 Data File : VY023330.D
 Acq On : 10 Sep 2025 14:02
 Operator : SY/MD
 Sample : Q3066-01
 Misc : 5.22g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 354

Quant Time: Sep 11 03:50:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090825S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 10 01:44:33 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	566690	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1084992	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	997278	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	405221	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	296029	59.448	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.900%
35) Dibromofluoromethane	7.634	113	354989	53.525	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.060%
50) Toluene-d8	10.109	98	1299730	51.899	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.800%
62) 4-Bromofluorobenzene	12.408	95	383685	48.684	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.360%
Target Compounds						
20) Methylene Chloride	4.616	84	24965	4.114	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091025\
Data File : VY023330.D
Acq On : 10 Sep 2025 14:02
Operator : SY/MD
Sample : Q3066-01
Misc : 5.22g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
354

Quant Time: Sep 11 03:50:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090825S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 10 01:44:33 2025
Response via : Initial Calibration

