

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091224\
 Data File : VY019520.D
 Acq On : 12 Sep 2024 11:22
 Operator : SY/MD
 Sample : VY0912SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0912SBL01

Quant Time: Sep 13 01:30:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	482360	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.615	114	832559	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	690395	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	309520	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	253262	56.702	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.400%
35) Dibromofluoromethane	7.634	113	254060	53.821	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.640%
50) Toluene-d8	10.109	98	936409	53.632	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	107.260%
62) 4-Bromofluorobenzene	12.407	95	326844	48.685	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	97.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091224\
Data File : VY019520.D
Acq On : 12 Sep 2024 11:22
Operator : SY/MD
Sample : VY0912SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0912SBL01

Quant Time: Sep 13 01:30:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

