

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091521\
 Data File : VY006017.D
 Acq On : 15 Sep 2021 10:45
 Operator : SY/MD
 Sample : VY0915SBL01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0915SBL01

Quant Time: Sep 16 01:21:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 15 05:01:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	108349	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	198374	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	209188	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	105405	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	81734	54.307	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.620%
35) Dibromofluoromethane	7.722	113	60387	49.010	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.020%
50) Toluene-d8	10.185	98	262275	49.944	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.880%
62) 4-Bromofluorobenzene	12.483	95	96243	53.980	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.960%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091521\
Data File : VY006017.D
Acq On : 15 Sep 2021 10:45
Operator : SY/MD
Sample : VY0915SBL01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0915SBL01

Quant Time: Sep 16 01:21:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 15 05:01:47 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.795 min Scan# 980
Delta R.T. 0.000 min
Lab File: VY006017.D
Acq: 15 Sep 2021 10:45




Instrument :
MSVOA_Y
ClientSampleId :
VY0915SBL01


Tgt Ion:168 Resp: 108349
Ion Ratio Lower Upper
168 100
99 54.6 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 54.307 ug/l
RT: 8.149 min Scan# 1038
Delta R.T. 0.000 min
Lab File: VY006017.D
Acq: 15 Sep 2021 10:45

Tgt Ion: 65 Resp: 81734
Ion Ratio Lower Upper
65 100
67 49.0 0.0 99.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.697 min Scan# 1128
Delta R.T. 0.000 min
Lab File: VY006017.D
Acq: 15 Sep 2021 10:45




Instrument :
MSVOA_Y
ClientSampleId :
VY0915SBL01


Tgt Ion:	114	Resp:	198374
Ion	Ratio	Lower	Upper
114	100		
63	22.0	0.0	43.2
88	16.1	0.0	31.8



#35
Dibromofluoromethane
Concen: 49.010 ug/l
RT: 7.722 min Scan# 968
Delta R.T. -0.006 min
Lab File: VY006017.D
Acq: 15 Sep 2021 10:45

Tgt Ion:	113	Resp:	60387
Ion	Ratio	Lower	Upper
113	100		
111	105.1	82.6	124.0
192	21.7	17.6	26.4









#50
Toluene-d8
Concen: 49.944 ug/l
RT: 10.185 min Scan# 1372
Delta R.T. 0.000 min
Lab File: VY006017.D
Acq: 15 Sep 2021 10:45




Instrument :
MSVOA_Y
ClientSampleId :
VY0915SBL01


Tgt Ion: 98 Resp: 262275
Ion Ratio Lower Upper
98 100
100 63.4 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 53.980 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY006017.D
Acq: 15 Sep 2021 10:45

Tgt Ion: 95 Resp: 96243
Ion Ratio Lower Upper
95 100
174 89.4 0.0 177.0
176 86.9 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 1586
Delta R.T. -0.006 min
Lab File: VY006017.D
Acq: 15 Sep 2021 10:45

Instrument :
MSVOA_Y
ClientSampleId :
VY0915SBL01

Tgt Ion:	117	Resp:	209188
Ion	Ratio	Lower	Upper
117	100		
82	60.1	43.0	64.4
119	33.2	25.4	38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY006017.D
Acq: 15 Sep 2021 10:45

Tgt Ion:	152	Resp:	105405
Ion	Ratio	Lower	Upper
152	100		
115	56.3	28.2	84.7
150	157.8	0.0	347.6

