

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092024\
 Data File : VY019642.D
 Acq On : 20 Sep 2024 13:49
 Operator : SY/MD
 Sample : P4098-03RE
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-04-091724RE

Quant Time: Sep 20 15:23:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	14509	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	25164	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	17026	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.341	152	3351	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	6549	48.746	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.500%
35) Dibromofluoromethane	7.628	113	6886	48.263	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.520%
50) Toluene-d8	10.103	98	27145	51.438	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.880%
62) 4-Bromofluorobenzene	12.402	95	5620	27.697	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	55.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092024\
Data File : VY019642.D
Acq On : 20 Sep 2024 13:49
Operator : SY/MD
Sample : P4098-03RE
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-04-091724RE

Quant Time: Sep 20 15:23:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

