

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092322\
 Data File : VY010642.D
 Acq On : 23 Sep 2022 12:39
 Operator : KP/MD
 Sample : N4723-19
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-29-6.5-7.0

Quant Time: Sep 24 01:20:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	3506	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	6707	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	7333	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	2131	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	2542	75.804	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	151.600%
35) Dibromofluoromethane	7.704	113	931	21.750	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	43.500%#
50) Toluene-d8	10.179	98	6227	40.027	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	80.060%
62) 4-Bromofluorobenzene	12.477	95	1831	30.113	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	60.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092322\
Data File : VY010642.D
Acq On : 23 Sep 2022 12:39
Operator : KP/MD
Sample : N4723-19
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-29-6.5-7.0

Quant Time: Sep 24 01:20:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

