

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092322\
 Data File : VY010662.D
 Acq On : 23 Sep 2022 20:26
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 24 01:25:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	70	0.00
2 T	Dichlorodifluoromethane	0.279	0.187	33.0#	60	0.00
3 P	Chloromethane	0.391	0.299	23.5	64	0.00
4 C	Vinyl Chloride	0.490	0.367	25.1#	60	0.00
5 T	Bromomethane	0.350	0.250	28.6#	60	0.00
6 T	Chloroethane	0.352	0.275	21.9	62	0.00
7 T	Trichlorofluoromethane	0.657	0.533	18.9	64	0.01
8 T	Diethyl Ether	0.225	0.202	10.2	68	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.395	0.340	13.9	65	0.01
10 T	Methyl Iodide	0.457	0.355	22.3	58	0.00
11 T	Tert butyl alcohol	0.049	0.050	-2.0	90	0.00
12 CM	1,1-Dichloroethene	0.353	0.280	20.7#	62	0.00
13 T	Acrolein	0.022	0.014	36.4#	51	0.00
14 T	Allyl chloride	0.532	0.476	10.5	69	0.00
15 T	Acrylonitrile	0.117	0.124	-6.0	80	0.00
16 T	Acetone	0.098	0.093	5.1	66	0.00
17 T	Carbon Disulfide	0.788	0.545	30.8#	59	0.01
18 T	Methyl Acetate	0.316	0.282	10.8	79	0.00
19 T	Methyl tert-butyl Ether	1.101	1.040	5.5	70	0.00
20 T	Methylene Chloride	0.785	0.505	35.7#	67	0.00
21 T	trans-1,2-Dichloroethene	0.397	0.328	17.4	64	0.00
22 T	Diisopropyl ether	1.257	1.287	-2.4	76	0.00
23 T	Vinyl Acetate	0.742	0.784	-5.7	75	0.00
24 P	1,1-Dichloroethane	0.754	0.718	4.8	71	0.00
25 T	2-Butanone	0.153	0.160	-4.6	78	0.00
26 T	2,2-Dichloropropane	0.725	0.633	12.7	66	0.00
27 T	cis-1,2-Dichloroethene	0.495	0.450	9.1	68	0.00
28 T	Bromochloromethane	0.301	0.403	-33.9#	78	0.00
29 T	Tetrahydrofuran	0.094	0.099	-5.3	79	0.00
30 C	Chloroform	0.816	0.795	2.6#	72	0.00
31 T	Cyclohexane	0.617	0.474	23.2	65	0.00
32 T	1,1,1-Trichloroethane	0.726	0.653	10.1	67	0.00
33 S	1,2-Dichloroethane-d4	0.478	0.522	-9.2	75	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	73	0.00
35 S	Dibromofluoromethane	0.319	0.299	6.3	72	0.00
36 T	1,1-Dichloropropene	0.356	0.295	17.1	65	0.00
37 T	Ethyl Acetate	0.205	0.211	-2.9	78	0.00
38 T	Carbon Tetrachloride	0.409	0.352	13.9	66	0.00
39 T	Methylcyclohexane	0.416	0.317	23.8	62	0.00
40 TM	Benzene	1.060	0.949	10.5	70	0.00
41 T	Methacrylonitrile	0.105	0.120	-14.3	85	0.00
42 TM	1,2-Dichloroethane	0.326	0.312	4.3	74	0.00
43 T	Isopropyl Acetate	0.376	0.381	-1.3	77	0.00
44 TM	Trichloroethene	0.303	0.259	14.5	67	0.00
45 C	1,2-Dichloropropane	0.274	0.261	4.7#	73	0.00
46 T	Dibromomethane	0.161	0.153	5.0	73	0.00
47 T	Bromodichloromethane	0.398	0.390	2.0	73	0.00
48 T	Methyl methacrylate	0.166	0.170	-2.4	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092322\
 Data File : VY010662.D
 Acq On : 23 Sep 2022 20:26
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 24 01:25:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	71	0.00
50 S	Toluene-d8	1.087	1.165	-7.2	71	0.00
51 T	4-Methyl-2-Pentanone	0.213	0.231	-8.5	81	0.00
52 CM	Toluene	0.694	0.619	10.8#	69	0.00
53 T	t-1,3-Dichloropropene	0.402	0.373	7.2	71	0.00
54 T	cis-1,3-Dichloropropene	0.448	0.415	7.4	70	0.00
55 T	1,1,2-Trichloroethane	0.237	0.234	1.3	75	0.00
56 T	Ethyl methacrylate	0.307	0.299	2.6	72	0.00
57 T	1,3-Dichloropropane	0.392	0.379	3.3	74	0.00
58 T	2-Chloroethyl Vinyl ether	0.155	0.166	-7.1	70	0.00
59 T	2-Hexanone	0.144	0.156	-8.3	79	0.00
60 T	Dibromochloromethane	0.290	0.287	1.0	74	0.00
61 T	1,2-Dibromoethane	0.222	0.207	6.8	71	0.00
62 S	4-Bromofluorobenzene	0.474	0.393	17.1	71	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	76	0.00
64 T	Tetrachloroethene	0.334	0.267	20.1	65	0.00
65 PM	Chlorobenzene	0.865	0.756	12.6	70	0.00
66 T	1,1,1,2-Tetrachloroethane	0.337	0.305	9.5	70	0.00
67 C	Ethyl Benzene	1.518	1.309	13.8#	69	0.00
68 T	m/p-Xylenes	0.588	0.505	14.1	69	0.00
69 T	o-Xylene	0.573	0.499	12.9	69	0.00
70 T	Styrene	0.974	0.866	11.1	70	0.00
71 P	Bromoform	0.209	0.195	6.7	73	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
73 T	Isopropylbenzene	3.090	2.614	15.4	68	0.00
74 T	N-amyl acetate	0.749	0.703	6.1	73	0.00
75 P	1,1,2,2-Tetrachloroethane	0.630	0.592	6.0	77	0.00
76 T	1,2,3-Trichloropropane	0.477	0.445	6.7	81	0.00
77 T	Bromobenzene	0.735	0.627	14.7	69	0.00
78 T	n-propylbenzene	3.681	3.162	14.1	69	0.00
79 T	2-Chlorotoluene	2.097	1.792	14.5	69	0.00
80 T	1,3,5-Trimethylbenzene	2.611	2.218	15.1	69	0.00
81 T	trans-1,4-Dichloro-2-butene	0.215	0.189	12.1	72	0.00
82 T	4-Chlorotoluene	2.166	1.864	13.9	70	0.00
83 T	tert-Butylbenzene	2.335	1.993	14.6	68	0.00
84 T	1,2,4-Trimethylbenzene	2.594	2.227	14.1	69	0.00
85 T	sec-Butylbenzene	3.420	2.941	14.0	69	0.00
86 T	p-Isopropyltoluene	2.858	2.468	13.6	68	0.00
87 T	1,3-Dichlorobenzene	1.492	1.274	14.6	69	0.00
88 T	1,4-Dichlorobenzene	1.477	1.270	14.0	69	0.00
89 T	n-Butylbenzene	2.624	2.244	14.5	68	0.00
90 T	Hexachloroethane	0.549	0.491	10.6	71	0.00
91 T	1,2-Dichlorobenzene	1.329	1.172	11.8	71	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.105	0.100	4.8	77	0.00
93 T	1,2,4-Trichlorobenzene	0.811	0.685	15.5	64	0.00
94 T	Hexachlorobutadiene	0.493	0.396	19.7	62	0.00
95 T	Naphthalene	1.634	1.456	10.9	67	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092322\
Data File : VY010662.D
Acq On : 23 Sep 2022 20:26
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Sep 24 01:25:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.706	0.604	14.4	65	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6