

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092424\
 Data File : VY019672.D
 Acq On : 24 Sep 2024 10:41
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 25 01:41:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	63	0.00
2 T	Dichlorodifluoromethane	0.316	0.236	25.3#	56	0.00
3 P	Chloromethane	0.414	0.364	12.1	58	0.00
4 C	Vinyl Chloride	0.447	0.456	-2.0#	64	0.00
5 T	Bromomethane	0.275	0.335	-21.8	74	0.00
6 T	Chloroethane	0.292	0.328	-12.3	69	0.00
7 T	Trichlorofluoromethane	0.746	0.795	-6.6	68	0.00
8 T	Diethyl Ether	0.250	0.257	-2.8	61	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.461	0.483	-4.8	66	0.00
10 T	Methyl Iodide	0.591	0.546	7.6	56	0.00
11 T	Tert butyl alcohol	0.037	0.041	-10.8	64	0.00
12 CM	1,1-Dichloroethene	0.444	0.449	-1.1#	63	0.00
13 T	Acrolein	0.020	0.009	55.0#	37#	0.00
14 T	Allyl chloride	0.796	0.770	3.3	59	0.00
15 T	Acrylonitrile	0.108	0.113	-4.6	61	0.00
16 T	Acetone	0.110	0.128	-16.4	65	0.00
17 T	Carbon Disulfide	1.199	1.127	6.0	55	0.00
18 T	Methyl Acetate	0.299	0.282	5.7	51	0.00
19 T	Methyl tert-butyl Ether	1.256	1.321	-5.2	63	0.00
20 T	Methylene Chloride	0.503	0.517	-2.8	65	0.00
21 T	trans-1,2-Dichloroethene	0.486	0.497	-2.3	62	0.00
22 T	Diisopropyl ether	1.634	1.662	-1.7	61	0.00
23 T	Vinyl Acetate	0.933	0.970	-4.0	63	0.00
24 P	1,1-Dichloroethane	0.917	0.956	-4.3	64	0.00
25 T	2-Butanone	0.157	0.167	-6.4	61	0.00
26 T	2,2-Dichloropropane	0.827	0.906	-9.6	68	0.00
27 T	cis-1,2-Dichloroethene	0.591	0.626	-5.9	64	0.00
28 T	Bromochloromethane	0.386	0.370	4.1	57	0.00
29 T	Tetrahydrofuran	0.095	0.095	0.0	58	0.00
30 C	Chloroform	0.936	1.011	-8.0#	66	0.00
31 T	Cyclohexane	0.829	0.778	6.2	61	0.00
32 T	1,1,1-Trichloroethane	0.839	0.909	-8.3	67	0.00
33 S	1,2-Dichloroethane-d4	0.463	0.450	2.8	58	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	62	0.00
35 S	Dibromofluoromethane	0.283	0.276	2.5	58	0.00
36 T	1,1-Dichloropropene	0.386	0.412	-6.7	65	0.00
37 T	Ethyl Acetate	0.198	0.207	-4.5	61	0.00
38 T	Carbon Tetrachloride	0.435	0.474	-9.0	67	0.00
39 T	Methylcyclohexane	0.513	0.534	-4.1	64	0.00
40 TM	Benzene	1.188	1.252	-5.4	64	0.00
41 T	Methacrylonitrile	0.115	0.115	0.0	54	0.00
42 TM	1,2-Dichloroethane	0.320	0.355	-10.9	66	0.00
43 T	Isopropyl Acetate	0.415	0.426	-2.7	61	0.00
44 TM	Trichloroethene	0.306	0.325	-6.2	64	0.00
45 C	1,2-Dichloropropane	0.290	0.301	-3.8#	63	0.00
46 T	Dibromomethane	0.159	0.174	-9.4	64	0.00
47 T	Bromodichloromethane	0.427	0.470	-10.1	66	0.00
48 T	Methyl methacrylate	0.196	0.195	0.5	57	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092424\
 Data File : VY019672.D
 Acq On : 24 Sep 2024 10:41
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 25 01:41:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	63	0.00
50 S	Toluene-d8	1.049	0.996	5.1	57	0.00
51 T	4-Methyl-2-Pentanone	0.209	0.215	-2.9	60	0.00
52 CM	Toluene	0.764	0.823	-7.7#	64	0.00
53 T	t-1,3-Dichloropropene	0.407	0.428	-5.2	62	0.00
54 T	cis-1,3-Dichloropropene	0.474	0.495	-4.4	62	0.00
55 T	1,1,2-Trichloroethane	0.219	0.234	-6.8	62	0.00
56 T	Ethyl methacrylate	0.333	0.348	-4.5	62	0.00
57 T	1,3-Dichloropropane	0.377	0.401	-6.4	63	0.00
58 T	2-Chloroethyl Vinyl ether	0.155	0.162	-4.5	63	0.00
59 T	2-Hexanone	0.153	0.158	-3.3	59	0.00
60 T	Dibromochloromethane	0.290	0.319	-10.0	65	0.00
61 T	1,2-Dibromoethane	0.204	0.215	-5.4	62	0.00
62 S	4-Bromofluorobenzene	0.403	0.362	10.2	56	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	61	0.00
64 T	Tetrachloroethene	0.310	0.323	-4.2	63	0.00
65 PM	Chlorobenzene	0.984	1.066	-8.3	65	0.00
66 T	1,1,1,2-Tetrachloroethane	0.341	0.372	-9.1	66	0.00
67 C	Ethyl Benzene	1.748	1.885	-7.8#	65	0.00
68 T	m/p-Xylenes	0.661	0.711	-7.6	65	0.00
69 T	o-Xylene	0.646	0.689	-6.7	64	0.00
70 T	Styrene	1.091	1.174	-7.6	65	0.00
71 P	Bromoform	0.195	0.212	-8.7	63	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	63	0.00
73 T	Isopropylbenzene	3.627	3.839	-5.8	67	0.00
74 T	N-amyl acetate	0.987	0.982	0.5	61	0.00
75 P	1,1,2,2-Tetrachloroethane	0.644	0.677	-5.1	64	0.00
76 T	1,2,3-Trichloropropane	0.469	0.512	-9.2	62	0.00
77 T	Bromobenzene	0.804	0.841	-4.6	65	0.00
78 T	n-propylbenzene	4.299	4.582	-6.6	67	0.00
79 T	2-Chlorotoluene	2.506	2.607	-4.0	65	0.00
80 T	1,3,5-Trimethylbenzene	2.949	3.103	-5.2	66	0.00
81 T	trans-1,4-Dichloro-2-butene	0.244	0.233	4.5	58	0.00
82 T	4-Chlorotoluene	2.576	2.692	-4.5	66	0.00
83 T	tert-Butylbenzene	2.670	2.884	-8.0	68	0.00
84 T	1,2,4-Trimethylbenzene	2.919	3.065	-5.0	66	0.00
85 T	sec-Butylbenzene	3.878	4.202	-8.4	68	0.00
86 T	p-Isopropyltoluene	3.173	3.469	-9.3	69	0.00
87 T	1,3-Dichlorobenzene	1.563	1.665	-6.5	67	0.00
88 T	1,4-Dichlorobenzene	1.551	1.641	-5.8	66	0.00
89 T	n-Butylbenzene	3.033	3.353	-10.6	69	0.00
90 T	Hexachloroethane	0.660	0.706	-7.0	67	0.00
91 T	1,2-Dichlorobenzene	1.395	1.490	-6.8	66	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.107	0.108	-0.9	61	0.00
93 T	1,2,4-Trichlorobenzene	0.848	0.922	-8.7	66	0.00
94 T	Hexachlorobutadiene	0.453	0.522	-15.2	74	0.00
95 T	Naphthalene	1.699	1.792	-5.5	63	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092424\
Data File : VY019672.D
Acq On : 24 Sep 2024 10:41
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Sep 25 01:41:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.726	0.787	-8.4	66	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6