

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092524\
 Data File : VY019695.D
 Acq On : 25 Sep 2024 15:14
 Operator : SY/MD
 Sample : P4172-06
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CHRT-24336

Quant Time: Sep 26 04:01:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.707	168	334913	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.615	114	639684	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	559239	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	203198	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	196556	63.380	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.760%
35) Dibromofluoromethane	7.634	113	203360	56.070	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.140%
50) Toluene-d8	10.109	98	747870	55.749	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	111.500%
62) 4-Bromofluorobenzene	12.407	95	240843	46.691	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	93.380%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092524\
 Data File : VY019695.D
 Acq On : 25 Sep 2024 15:14
 Operator : SY/MD
 Sample : P4172-06
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CHRT-24336

Quant Time: Sep 26 04:01:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

