

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092524\
 Data File : VY019699.D
 Acq On : 25 Sep 2024 16:48
 Operator : SY/MD
 Sample : P4167-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PL-03-09242024

Quant Time: Sep 26 04:02:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	326625	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	621399	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	558571	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	216314	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	193865	64.099	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	128.200%	
35) Dibromofluoromethane	7.634	113	198408	56.314	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	112.620%	
50) Toluene-d8	10.109	98	736253	56.498	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	113.000%	
62) 4-Bromofluorobenzene	12.408	95	242692	48.434	ug/l	0.00
Spiked Amount	50.000	Range 29 - 146	Recovery	=	96.860%	
Target Compounds						
16) Acetone	3.873	43	6390	8.868	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092524\
Data File : VY019699.D
Acq On : 25 Sep 2024 16:48
Operator : SY/MD
Sample : P4167-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PL-03-09242024

Quant Time: Sep 26 04:02:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

