

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092524\
 Data File : VY019689.D
 Acq On : 25 Sep 2024 11:19
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 26 03:55:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	59	0.00
2 T	Dichlorodifluoromethane	0.316	0.242	23.4	54	0.00
3 P	Chloromethane	0.414	0.372	10.1	56	0.00
4 C	Vinyl Chloride	0.447	0.464	-3.8#	61	0.00
5 T	Bromomethane	0.275	0.336	-22.2	70	0.00
6 T	Chloroethane	0.292	0.341	-16.8	68	0.00
7 T	Trichlorofluoromethane	0.746	0.824	-10.5	66	0.00
8 T	Diethyl Ether	0.250	0.259	-3.6	58	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.461	0.493	-6.9	64	0.00
10 T	Methyl Iodide	0.591	0.559	5.4	54	0.00
11 T	Tert butyl alcohol	0.037	0.039	-5.4	57	0.00
12 CM	1,1-Dichloroethene	0.444	0.460	-3.6#	61	0.00
13 T	Acrolein	0.020	0.012	40.0#	44#	0.00
14 T	Allyl chloride	0.796	0.773	2.9	56	0.00
15 T	Acrylonitrile	0.108	0.113	-4.6	58	0.00
16 T	Acetone	0.110	0.121	-10.0	58	0.00
17 T	Carbon Disulfide	1.199	1.139	5.0	52	0.00
18 T	Methyl Acetate	0.299	0.287	4.0	49#	0.00
19 T	Methyl tert-butyl Ether	1.256	1.342	-6.8	60	0.00
20 T	Methylene Chloride	0.503	0.516	-2.6	61	0.00
21 T	trans-1,2-Dichloroethene	0.486	0.511	-5.1	60	0.00
22 T	Diisopropyl ether	1.634	1.702	-4.2	59	0.00
23 T	Vinyl Acetate	0.933	0.967	-3.6	59	0.00
24 P	1,1-Dichloroethane	0.917	0.982	-7.1	61	0.00
25 T	2-Butanone	0.157	0.160	-1.9	55	0.00
26 T	2,2-Dichloropropane	0.827	0.915	-10.6	65	0.00
27 T	cis-1,2-Dichloroethene	0.591	0.640	-8.3	62	0.00
28 T	Bromochloromethane	0.386	0.428	-10.9	62	0.00
29 T	Tetrahydrofuran	0.095	0.095	0.0	55	0.00
30 C	Chloroform	0.936	1.041	-11.2#	64	0.00
31 T	Cyclohexane	0.829	0.783	5.5	58	0.00
32 T	1,1,1-Trichloroethane	0.839	0.935	-11.4	65	0.00
33 S	1,2-Dichloroethane-d4	0.463	0.499	-7.8	61	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	58	0.00
35 S	Dibromofluoromethane	0.283	0.307	-8.5	62	0.00
36 T	1,1-Dichloropropene	0.386	0.417	-8.0	62	0.00
37 T	Ethyl Acetate	0.198	0.195	1.5	54	0.00
38 T	Carbon Tetrachloride	0.435	0.479	-10.1	64	0.00
39 T	Methylcyclohexane	0.513	0.525	-2.3	59	0.00
40 TM	Benzene	1.188	1.265	-6.5	61	0.00
41 T	Methacrylonitrile	0.115	0.126	-9.6	56	0.00
42 TM	1,2-Dichloroethane	0.320	0.358	-11.9	63	0.00
43 T	Isopropyl Acetate	0.415	0.416	-0.2	56	0.00
44 TM	Trichloroethene	0.306	0.328	-7.2	61	0.00
45 C	1,2-Dichloropropane	0.290	0.309	-6.6#	62	0.00
46 T	Dibromomethane	0.159	0.172	-8.2	60	0.00
47 T	Bromodichloromethane	0.427	0.471	-10.3	63	0.00
48 T	Methyl methacrylate	0.196	0.191	2.6	53	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092524\
 Data File : VY019689.D
 Acq On : 25 Sep 2024 11:19
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 26 03:55:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	58	0.00
50 S	Toluene-d8	1.049	1.106	-5.4	60	0.00
51 T	4-Methyl-2-Pentanone	0.209	0.211	-1.0	56	0.00
52 CM	Toluene	0.764	0.831	-8.8#	61	0.00
53 T	t-1,3-Dichloropropene	0.407	0.429	-5.4	59	0.00
54 T	cis-1,3-Dichloropropene	0.474	0.497	-4.9	59	0.00
55 T	1,1,2-Trichloroethane	0.219	0.235	-7.3	59	0.00
56 T	Ethyl methacrylate	0.333	0.349	-4.8	59	0.00
57 T	1,3-Dichloropropane	0.377	0.403	-6.9	60	0.00
58 T	2-Chloroethyl Vinyl ether	0.155	0.164	-5.8	60	0.00
59 T	2-Hexanone	0.153	0.154	-0.7	55	0.00
60 T	Dibromochloromethane	0.290	0.318	-9.7	61	0.00
61 T	1,2-Dibromoethane	0.204	0.216	-5.9	59	0.00
62 S	4-Bromofluorobenzene	0.403	0.406	-0.7	59	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	57	0.00
64 T	Tetrachloroethene	0.310	0.332	-7.1	61	0.00
65 PM	Chlorobenzene	0.984	1.088	-10.6	63	0.00
66 T	1,1,1,2-Tetrachloroethane	0.341	0.382	-12.0	64	0.00
67 C	Ethyl Benzene	1.748	1.932	-10.5#	63	0.00
68 T	m/p-Xylenes	0.661	0.727	-10.0	62	0.00
69 T	o-Xylene	0.646	0.701	-8.5	61	0.00
70 T	Styrene	1.091	1.190	-9.1	61	0.00
71 P	Bromoform	0.195	0.212	-8.7	59	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	59	0.00
73 T	Isopropylbenzene	3.627	3.908	-7.7	64	0.00
74 T	N-amyl acetate	0.987	0.990	-0.3	58	0.00
75 P	1,1,2,2-Tetrachloroethane	0.644	0.673	-4.5	60	0.00
76 T	1,2,3-Trichloropropane	0.469	0.430	8.3	49#	0.00
77 T	Bromobenzene	0.804	0.851	-5.8	62	0.00
78 T	n-propylbenzene	4.299	4.607	-7.2	63	0.00
79 T	2-Chlorotoluene	2.506	2.634	-5.1	62	0.00
80 T	1,3,5-Trimethylbenzene	2.949	3.155	-7.0	63	0.00
81 T	trans-1,4-Dichloro-2-butene	0.244	0.230	5.7	54	0.00
82 T	4-Chlorotoluene	2.576	2.701	-4.9	62	0.00
83 T	tert-Butylbenzene	2.670	2.900	-8.6	64	0.00
84 T	1,2,4-Trimethylbenzene	2.919	3.142	-7.6	63	0.00
85 T	sec-Butylbenzene	3.878	4.174	-7.6	64	0.00
86 T	p-Isopropyltoluene	3.173	3.481	-9.7	65	0.00
87 T	1,3-Dichlorobenzene	1.563	1.705	-9.1	64	0.00
88 T	1,4-Dichlorobenzene	1.551	1.678	-8.2	63	0.00
89 T	n-Butylbenzene	3.033	3.335	-10.0	65	0.00
90 T	Hexachloroethane	0.660	0.700	-6.1	63	0.00
91 T	1,2-Dichlorobenzene	1.395	1.508	-8.1	63	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.107	0.109	-1.9	59	0.00
93 T	1,2,4-Trichlorobenzene	0.848	0.929	-9.6	63	0.00
94 T	Hexachlorobutadiene	0.453	0.493	-8.8	65	0.00
95 T	Naphthalene	1.699	1.916	-12.8	63	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092524\
Data File : VY019689.D
Acq On : 25 Sep 2024 11:19
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Sep 26 03:55:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.726	0.785	-8.1	62	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6