

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092525\
 Data File : VY023368.D
 Acq On : 25 Sep 2025 12:58
 Operator : SY/MD
 Sample : Q3189-01
 Misc : 5.25g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11989

Quant Time: Sep 25 16:29:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090825S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 10 01:44:33 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	656846	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	1188101	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	1165416	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	549362	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	321066	55.627	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.260%
35) Dibromofluoromethane	7.634	113	379471	52.251	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.500%
50) Toluene-d8	10.103	98	1392736	50.786	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.580%
62) 4-Bromofluorobenzene	12.401	95	509202	59.003	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	118.000%
Target Compounds						
20) Methylene Chloride	4.610	84	43279	6.153	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092525\
Data File : VY023368.D
Acq On : 25 Sep 2025 12:58
Operator : SY/MD
Sample : Q3189-01
Misc : 5.25g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11989

Quant Time: Sep 25 16:29:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090825S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 10 01:44:33 2025
Response via : Initial Calibration

