

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092622\
 Data File : VY010667.D
 Acq On : 26 Sep 2022 13:49
 Operator : KP/MD
 Sample : N4752-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-07-092122

Quant Time: Sep 26 14:53:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	151276	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	274039	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	259658	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	113115	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	95311	65.872	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.740%
35) Dibromofluoromethane	7.716	113	86669	49.556	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.120%
50) Toluene-d8	10.179	98	330653	51.084	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.160%
62) 4-Bromofluorobenzene	12.483	95	108152	44.805	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092622\
Data File : VY010667.D
Acq On : 26 Sep 2022 13:49
Operator : KP/MD
Sample : N4752-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-07-092122

Quant Time: Sep 26 14:53:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

