

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092622\
 Data File : VY010672.D
 Acq On : 26 Sep 2022 15:42
 Operator : KP/MD
 Sample : N4762-02RE
 Misc : 5.62g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 D22014-(17.5-18)RE

Quant Time: Sep 27 06:27:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	174217	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	294662	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	220308	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	60330	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	94904	56.953	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.900%
35) Dibromofluoromethane	7.715	113	93561	49.752	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.500%
50) Toluene-d8	10.178	98	278321	40.668	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	81.340%
62) 4-Bromofluorobenzene	12.477	95	64063	23.400	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	46.800%
Target Compounds						
16) Acetone	3.942	43	9756	28.710	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092622\
Data File : VY010672.D
Acq On : 26 Sep 2022 15:42
Operator : KP/MD
Sample : N4762-02RE
Misc : 5.62g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
D22014-(17.5-18)RE

Quant Time: Sep 27 06:27:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

