

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092622\
 Data File : VY010681.D
 Acq On : 26 Sep 2022 19:06
 Operator : KP/MD
 Sample : N4780-11
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP-2

Quant Time: Sep 27 06:28:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	218154	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	369298	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	341885	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	169478	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	111257	53.320	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.640%
35) Dibromofluoromethane	7.716	113	110689	46.965	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.920%
50) Toluene-d8	10.179	98	374133	43.392	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	86.780%
62) 4-Bromofluorobenzene	12.477	95	132261	40.395	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092622\
Data File : VY010681.D
Acq On : 26 Sep 2022 19:06
Operator : KP/MD
Sample : N4780-11
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SP-2

Quant Time: Sep 27 06:28:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

