

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092622\
 Data File : VY010682.D
 Acq On : 26 Sep 2022 19:29
 Operator : KP/MD
 Sample : N4780-31
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP-4

Quant Time: Sep 27 06:28:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	204744	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	348809	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	321325	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	155094	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	109908	56.123	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.240%
35) Dibromofluoromethane	7.716	113	109524	49.200	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.400%
50) Toluene-d8	10.179	98	366840	44.927	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.860%
62) 4-Bromofluorobenzene	12.483	95	128881	41.765	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092622\
Data File : VY010682.D
Acq On : 26 Sep 2022 19:29
Operator : KP/MD
Sample : N4780-31
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SP-4

Quant Time: Sep 27 06:28:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

