

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092722\
 Data File : VY010694.D
 Acq On : 27 Sep 2022 13:44
 Operator : KP/MD
 Sample : N4751-02
 Misc : 4.65g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PE-2

Quant Time: Sep 28 01:25:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	40799	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	63076	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	53122	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	19388	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	21432	54.921	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.840%	
35) Dibromofluoromethane	7.716	113	21503	53.417	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.840%	
50) Toluene-d8	10.179	98	63295	43.010	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 86.020%	
62) 4-Bromofluorobenzene	12.477	95	16992	29.677	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 59.360%	
Target Compounds						Qvalue
16) Acetone	3.942	43	2109	26.502	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092722\
Data File : VY010694.D
Acq On : 27 Sep 2022 13:44
Operator : KP/MD
Sample : N4751-02
Misc : 4.65g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PE-2

Quant Time: Sep 28 01:25:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

