

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092722\
 Data File : VY010697.D
 Acq On : 27 Sep 2022 14:52
 Operator : KP/MD
 Sample : N4751-05
 Misc : 10.16g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PE-5

Quant Time: Sep 28 01:25:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	155058	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	215374	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	150188	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	50556	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	35209	23.740	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	47.480%#
35) Dibromofluoromethane	7.710	113	48768	35.480	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	70.960%
50) Toluene-d8	10.179	98	202806	40.552	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	81.100%
62) 4-Bromofluorobenzene	12.477	95	45246	22.515	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	45.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092722\
Data File : VY010697.D
Acq On : 27 Sep 2022 14:52
Operator : KP/MD
Sample : N4751-05
Misc : 10.16g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PE-5

Quant Time: Sep 28 01:25:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

