

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092722\
 Data File : VY010702.D
 Acq On : 27 Sep 2022 16:45
 Operator : KP/MD
 Sample : N4808-01
 Misc : 6.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 D22022-6-6.5

Quant Time: Sep 28 01:26:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	170878	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	293870	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	274035	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	138094	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	94572	57.863	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.720%
35) Dibromofluoromethane	7.716	113	94640	50.462	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.920%
50) Toluene-d8	10.179	98	309799	45.027	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	90.060%
62) 4-Bromofluorobenzene	12.483	95	111343	42.900	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	85.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092722\
Data File : VY010702.D
Acq On : 27 Sep 2022 16:45
Operator : KP/MD
Sample : N4808-01
Misc : 6.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
D22022-6-6.5

Quant Time: Sep 28 01:26:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

