



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092722\
Data File : VY010685.D
Acq On : 27 Sep 2022 09:03
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Title : SW846 8260
Last Update : Tue Sep 20 01:02:44 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.6	16226	PASS
75	95	30	60	52.3	43208	PASS
95	95	100	100	100.0	82592	PASS
96	95	5	9	6.7	5542	PASS
173	174	0.00	2	1.1	796	PASS
174	95	50	100	84.3	69632	PASS
175	174	5	9	7.2	5031	PASS
176	174	95	101	95.4	66443	PASS
177	176	5	9	7.3	4858	PASS