

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092922\
 Data File : VY010735.D
 Acq On : 29 Sep 2022 16:45
 Operator : KP/MD
 Sample : N4868-02
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ROLL-OFF-349-ETGI

Quant Time: Sep 30 08:23:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	232348	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	379389	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	347191	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	167571	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	121263	54.565	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.140%	
35) Dibromofluoromethane	7.710	113	125502	51.833	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.660%	
50) Toluene-d8	10.179	98	406799	45.744	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 91.480%	
62) 4-Bromofluorobenzene	12.477	95	150025	44.899	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 89.800%	
Target Compounds						
16) Acetone	3.948	43	5533	12.209	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092922\
Data File : VY010735.D
Acq On : 29 Sep 2022 16:45
Operator : KP/MD
Sample : N4868-02
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ROLL-OFF-349-ETGI

Quant Time: Sep 30 08:23:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

