

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092922\
 Data File : VY010736.D
 Acq On : 29 Sep 2022 17:07
 Operator : KP/MD
 Sample : N4898-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CRUSHED-MASONRY-PILE-11

Quant Time: Sep 30 08:23:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	250912	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	411699	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	373958	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	184183	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	130324	54.304	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.600%
35) Dibromofluoromethane	7.710	113	15476	5.890	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	11.780%#
50) Toluene-d8	10.179	98	443068	45.901	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.800%
62) 4-Bromofluorobenzene	12.477	95	163149	45.001	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092922\
Data File : VY010736.D
Acq On : 29 Sep 2022 17:07
Operator : KP/MD
Sample : N4898-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CRUSHED-MASONRY-PILE-11

Quant Time: Sep 30 08:23:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

