

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092922\
 Data File : VY010738.D
 Acq On : 29 Sep 2022 17:53
 Operator : KP/MD
 Sample : N4896-01
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CRUSHED-MASONRY-PILE-9

Quant Time: Sep 30 08:24:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 01:02:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	176732	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	294296	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	259332	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	109372	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	99416	58.812	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.620%
35) Dibromofluoromethane	7.716	113	2659	1.416	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	2.840%#
50) Toluene-d8	10.179	98	299232	43.539	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	87.080%
62) 4-Bromofluorobenzene	12.477	95	106490	40.842	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.680%
Target Compounds						
16) Acetone	3.948	43	3531	10.243	ug/l	91
70) Styrene	12.044	104	5359	1.061	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY092922\
Data File : VY010738.D
Acq On : 29 Sep 2022 17:53
Operator : KP/MD
Sample : N4896-01
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CRUSHED-MASONRY-PILE-9

Quant Time: Sep 30 08:24:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091922S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 01:02:44 2022
Response via : Initial Calibration

