

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100124\
 Data File : VY019744.D
 Acq On : 01 Oct 2024 16:24
 Operator : SY/MD
 Sample : P4212-01RE
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TEST-PIT-NORTHRE

Quant Time: Oct 02 03:15:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	6570	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.615	114	10915	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	5432	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	646	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	3259	53.570	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.140%
35) Dibromofluoromethane	7.634	113	2730	44.113	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.220%
50) Toluene-d8	10.109	98	10853	47.413	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.820%
62) 4-Bromofluorobenzene	12.395	95	1056	11.998	ug/l	-0.01
Spiked Amount	50.000	Range	29 - 146	Recovery	=	24.000%#
Target Compounds						
16) Acetone	3.866	43	1866	128.748	ug/l	# 67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100124\
Data File : VY019744.D
Acq On : 01 Oct 2024 16:24
Operator : SY/MD
Sample : P4212-01RE
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TEST-PIT-NORTHRE

Quant Time: Oct 02 03:15:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

