

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100124\
 Data File : VY019747.D
 Acq On : 01 Oct 2024 17:34
 Operator : SY/MD
 Sample : P4222-03
 Misc : 4.52g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-5-VOC

Quant Time: Oct 02 03:17:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	299813	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.615	114	567094	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	500826	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	191207	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	173905	62.641	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 125.280%	
35) Dibromofluoromethane	7.634	113	178970	55.662	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 111.320%	
50) Toluene-d8	10.103	98	671976	56.503	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 113.000%	
62) 4-Bromofluorobenzene	12.401	95	222803	48.723	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 97.440%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100124\
Data File : VY019747.D
Acq On : 01 Oct 2024 17:34
Operator : SY/MD
Sample : P4222-03
Misc : 4.52g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-5-VOC

Quant Time: Oct 02 03:17:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

