

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100124\
 Data File : VY019749.D
 Acq On : 01 Oct 2024 18:21
 Operator : SY/MD
 Sample : P4234-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-02-093024

Quant Time: Oct 02 03:18:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	294830	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	562697	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	498509	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	182833	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	173599	63.588	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 127.180%	
35) Dibromofluoromethane	7.634	113	176378	55.284	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.560%	
50) Toluene-d8	10.103	98	666164	56.452	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 112.900%	
62) 4-Bromofluorobenzene	12.401	95	218834	48.229	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 96.460%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100124\
Data File : VY019749.D
Acq On : 01 Oct 2024 18:21
Operator : SY/MD
Sample : P4234-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-02-093024

Quant Time: Oct 02 03:18:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

