

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100124\
 Data File : VY019736.D
 Acq On : 01 Oct 2024 11:20
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 01 17:07:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	53	0.00
2 T	Dichlorodifluoromethane	0.316	0.224	29.1#	45#	0.00
3 P	Chloromethane	0.414	0.364	12.1	49#	0.00
4 C	Vinyl Chloride	0.447	0.435	2.7#	52	0.00
5 T	Bromomethane	0.275	0.299	-8.7	56	0.00
6 T	Chloroethane	0.292	0.315	-7.9	56	0.00
7 T	Trichlorofluoromethane	0.746	0.757	-1.5	55	0.00
8 T	Diethyl Ether	0.250	0.252	-0.8	51	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.461	0.467	-1.3	54	0.00
10 T	Methyl Iodide	0.591	0.495	16.2	43#	0.00
11 T	Tert butyl alcohol	0.037	0.046	-24.3	61	0.00
12 CM	1,1-Dichloroethene	0.444	0.432	2.7#	51	0.00
13 T	Acrolein	0.020	0.009	55.0#	31#	0.00
14 T	Allyl chloride	0.796	0.802	-0.8	52	0.00
15 T	Acrylonitrile	0.108	0.115	-6.5	53	0.00
16 T	Acetone	0.110	0.128	-16.4	55	0.00
17 T	Carbon Disulfide	1.199	1.089	9.2	45#	0.00
18 T	Methyl Acetate	0.299	0.305	-2.0	47#	0.00
19 T	Methyl tert-butyl Ether	1.256	1.302	-3.7	53	0.00
20 T	Methylene Chloride	0.503	0.493	2.0	52	0.00
21 T	trans-1,2-Dichloroethene	0.486	0.484	0.4	51	0.01
22 T	Diisopropyl ether	1.634	1.732	-6.0	54	0.00
23 T	Vinyl Acetate	0.933	0.981	-5.1	54	0.00
24 P	1,1-Dichloroethane	0.917	0.963	-5.0	54	0.00
25 T	2-Butanone	0.157	0.170	-8.3	53	0.00
26 T	2,2-Dichloropropane	0.827	0.911	-10.2	58	0.00
27 T	cis-1,2-Dichloroethene	0.591	0.614	-3.9	53	0.00
28 T	Bromochloromethane	0.386	0.471	-22.0	61	0.00
29 T	Tetrahydrofuran	0.095	0.098	-3.2	50	0.00
30 C	Chloroform	0.936	1.004	-7.3#	55	0.00
31 T	Cyclohexane	0.829	0.781	5.8	52	0.00
32 T	1,1,1-Trichloroethane	0.839	0.897	-6.9	56	0.00
33 S	1,2-Dichloroethane-d4	0.463	0.535	-15.6	59	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	52	0.00
35 S	Dibromofluoromethane	0.283	0.318	-12.4	57	0.00
36 T	1,1-Dichloropropene	0.386	0.400	-3.6	53	0.00
37 T	Ethyl Acetate	0.198	0.205	-3.5	51	0.00
38 T	Carbon Tetrachloride	0.435	0.456	-4.8	55	0.00
39 T	Methylcyclohexane	0.513	0.506	1.4	51	0.00
40 TM	Benzene	1.188	1.231	-3.6	53	0.00
41 T	Methacrylonitrile	0.115	0.126	-9.6	50	0.00
42 TM	1,2-Dichloroethane	0.320	0.348	-8.7	55	0.00
43 T	Isopropyl Acetate	0.415	0.428	-3.1	52	0.00
44 TM	Trichloroethene	0.306	0.310	-1.3	52	0.00
45 C	1,2-Dichloropropane	0.290	0.308	-6.2#	55	0.00
46 T	Dibromomethane	0.159	0.166	-4.4	52	0.00
47 T	Bromodichloromethane	0.427	0.461	-8.0	55	0.00
48 T	Methyl methacrylate	0.196	0.196	0.0	49#	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100124\
 Data File : VY019736.D
 Acq On : 01 Oct 2024 11:20
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 01 17:07:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	50	0.00
50 S	Toluene-d8	1.049	1.147	-9.3	55	0.00
51 T	4-Methyl-2-Pentanone	0.209	0.217	-3.8	51	0.00
52 CM	Toluene	0.764	0.798	-4.5#	53	0.00
53 T	t-1,3-Dichloropropene	0.407	0.424	-4.2	52	0.00
54 T	cis-1,3-Dichloropropene	0.474	0.496	-4.6	53	0.00
55 T	1,1,2-Trichloroethane	0.219	0.228	-4.1	52	0.00
56 T	Ethyl methacrylate	0.333	0.342	-2.7	52	0.00
57 T	1,3-Dichloropropane	0.377	0.391	-3.7	52	0.00
58 T	2-Chloroethyl Vinyl ether	0.155	0.182	-17.4	60	0.00
59 T	2-Hexanone	0.153	0.157	-2.6	50	0.00
60 T	Dibromochloromethane	0.290	0.304	-4.8	52	0.00
61 T	1,2-Dibromoethane	0.204	0.203	0.5	50	0.00
62 S	4-Bromofluorobenzene	0.403	0.423	-5.0	55	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	51	0.00
64 T	Tetrachloroethene	0.310	0.310	0.0	51	-0.01
65 PM	Chlorobenzene	0.984	1.041	-5.8	53	0.00
66 T	1,1,1,2-Tetrachloroethane	0.341	0.365	-7.0	54	0.00
67 C	Ethyl Benzene	1.748	1.877	-7.4#	54	-0.01
68 T	m/p-Xylenes	0.661	0.695	-5.1	53	-0.01
69 T	o-Xylene	0.646	0.669	-3.6	52	0.00
70 T	Styrene	1.091	1.140	-4.5	52	0.00
71 P	Bromoform	0.195	0.199	-2.1	50#	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	51	0.00
73 T	Isopropylbenzene	3.627	3.821	-5.3	54	0.00
74 T	N-amyl acetate	0.987	0.992	-0.5	50#	0.00
75 P	1,1,2,2-Tetrachloroethane	0.644	0.671	-4.2	52	-0.01
76 T	1,2,3-Trichloropropane	0.469	0.488	-4.1	48#	-0.01
77 T	Bromobenzene	0.804	0.823	-2.4	52	-0.01
78 T	n-propylbenzene	4.299	4.595	-6.9	55	0.00
79 T	2-Chlorotoluene	2.506	2.653	-5.9	54	0.00
80 T	1,3,5-Trimethylbenzene	2.949	3.073	-4.2	53	0.00
81 T	trans-1,4-Dichloro-2-butene	0.244	0.237	2.9	48#	0.00
82 T	4-Chlorotoluene	2.576	2.700	-4.8	54	0.00
83 T	tert-Butylbenzene	2.670	2.868	-7.4	55	0.00
84 T	1,2,4-Trimethylbenzene	2.919	3.027	-3.7	53	-0.01
85 T	sec-Butylbenzene	3.878	4.114	-6.1	54	0.00
86 T	p-Isopropyltoluene	3.173	3.345	-5.4	54	0.00
87 T	1,3-Dichlorobenzene	1.563	1.617	-3.5	53	-0.01
88 T	1,4-Dichlorobenzene	1.551	1.593	-2.7	52	-0.01
89 T	n-Butylbenzene	3.033	3.227	-6.4	54	-0.01
90 T	Hexachloroethane	0.660	0.687	-4.1	53	-0.01
91 T	1,2-Dichlorobenzene	1.395	1.436	-2.9	52	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.107	0.107	0.0	50#	0.00
93 T	1,2,4-Trichlorobenzene	0.848	0.812	4.2	48#	0.00
94 T	Hexachlorobutadiene	0.453	0.446	1.5	51	0.00
95 T	Naphthalene	1.699	1.550	8.8	44#	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100124\
Data File : VY019736.D
Acq On : 01 Oct 2024 11:20
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 01 17:07:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.726	0.667	8.1	45#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6